

METHOD FOR CALIBRATING SPECTROPHOTOMETRIC

This is a Continuation-In-Part Application of a U.S. Application No. 09/773,495 filed February 2, 2001, which is a continuation-in-part of a U.S. Application No. 09/697,679 filed October 27, 2000, which is a continuation-in-
5 part of U.S. Application No. 09/447,215 filed November 23, 1999.

FIELD OF INVENTION

This invention is in the field of spectrophotometric determinations of concentrations of analytes in samples. The invention further relates to
10 methods of calibrating spectrophotometers. Particularly, the method relates to the calibration of spectrophotometric apparatus designed to measure interferents in serum and plasma.

BACKGROUND OF INVENTION

15 Clinical laboratory tests are routinely performed on serum or plasma of whole blood. In a routine assay, red blood cells are separated from plasma by centrifugation. Red blood cells and various plasma proteins may also be separated from serum by clotting prior to centrifugation. Hemoglobin (Hb), light-scattering substances like lipid particles, bile pigments bilirubin (BR) and
20 biliverdin (BV) are typical substances which will interfere with and affect spectrophotometric and other blood analytical measurements and are therefore referred to as interferents. The presence of such interferents affects the ability to perform tests on the serum or plasma and as such can be said to compromise specimen integrity.

25

Visual inspection can be used to determine the presence of interferents in serum and plasma but such a method relies on the experience and knowledge of the observer and is therefore unreliable. The use of an apparatus or instrument to measure interferents in serum and plasma i.e.,
30 assess specimen integrity, is a substitute for visual inspection and the

interferents may be regarded as analytes with respect to the apparatus. Measurement of interferents are taught in WO 9838961 and WO 9839634. Because quantitative results from the determination of the concentration of such interferents are reported based on specific calibration algorithms, there
5 is a need to calibrate and to monitor calibration performance daily.

Unlike many blood analytical apparatus, calibration of reagentless spectrophotometric apparatus used to measure the concentration of analytes or interferents in a serum or plasma sample is a cumbersome time intensive
10 exercise. Each apparatus used for the purposes of determining the concentration of interferents must be calibrated according to procedures known in the art, for example, the process described herein, in the section titled "Primary Calibration," and over the lifetime of an apparatus can amount to a considerable amount of time and cost. Furthermore, in settings where a
15 large number of apparatus is needed to perform multiple sample measurements (such as blood banks for example) the time required for calibration can become a real burden on the efficiency of the of the quality control process.

20 Martinek (J. Amer. Med. Technol., July-August 1978, p. 210 – 216) teaches a method of photometric correction, involving liquid absorbance standards to correct one spectrophotometer to match another using a slope and bias correction. This method may also be used for test methods that require reagents.

25

U.S. 4,866,644 teaches a method of calibrating a second apparatus to produce results for a test sample, as if the sample was tested on a first apparatus. The method combines photometric correction with a mathematical process that computes a waveshift for each index point. The
30 waveshifts are derived from the assessment of readings determined for a

plurality of samples on the two apparatus. The waveshifts are applied as corrections to an existing wavelength calibration table of the second apparatus in order to make the second apparatus behave in a manner similar to the first apparatus. In U.S. 4,866,644, the same wavelengths are assigned
5 to the same corresponding index points in every instrument. Therefore, there is no derivation of a new wavelength calibration table of the second instrument, and the waveshift correction is applied to each measurement as it is determined on the second instrument

10 WO 94/08225 discloses a method involving the modification of the constants of a primary calibration algorithm of a second or recalibrated apparatus, to yield results consistent with a first apparatus that is in control. A limitation of this method is that the number of samples required must be at least one more than the number of terms used in the primary calibration
15 equation, because a mathematical system of "simultaneous equations" is used to generate a new constant for each term in the primary calibration algorithm. Furthermore, a predicted dependant variable, such as a chemical or physical property, of a calibrator is required to generate the new constants.

20 WO 97/47942 teaches a method for a second apparatus to produce results for a test sample, as if the sample was tested on the first apparatus involving testing a set of stable samples, whose absorbance spectra mimic that of the analytes, on both the first and a second apparatus, and predicting the analyte concentrations after applying a primary calibration algorithm. This
25 method requires a predicted dependant variable, for example a chemical or physical property, of the calibrator. Analyte results predicted by both apparatus are used to perform a slope and bias correction of each analyte prediction on a test sample. The calibration set requires the property of having an absorbance spectra similar to the analyte.

There is a need for a method to simply and accurately calibrate a second apparatus, and to recalibrate a first or second apparatus that is no longer in control.

5 It is an object of the invention to overcome disadvantages of the prior art.

The above object is met by the combinations of features of the main claims, the sub-claims disclose further advantageous embodiments of the
10 invention.

SUMMARY OF INVENTION

This invention is in the field of spectrophotometric determination of concentrations of analytes in samples. The invention further relates to
15 methods of calibrating spectrophotometers. The method may be used for the calibration of spectrophotometric apparatus designed to measure interferences in serum and plasma. The invention also relates to a method of transferring calibration algorithms from a first apparatus to a second apparatus.

20 The present inventor has found that for a given analyte, a "Primary Calibration Algorithm" developed for a "First apparatus" can be transferred onto a "Second Apparatus". Therefore, the Second Apparatus need not be subjected to the cumbersome, time intensive Primary Calibration process.

25 In one aspect of the invention, the First Apparatus that is known to be "In Control" is used to assign absorbance values to a "Set of Calibrators" from a batch or lot, and any Second Apparatus can be calibrated rapidly by a process of "Calibration Algorithm Transfer," and the concentration of an analyte in a sample determined by applying the "Primary Calibration
30 Algorithm" to a corrected interpolated absorbance measurement of the

sample. Therefore, the present invention provides a method for calibrating a Second Apparatus using a Set of Calibrators with absorbances assigned by the First Apparatus.

5 In yet a further aspect of the invention a method for adjusting the absorbance of sample obtained on a second apparatus to normalize it with that of a first apparatus that is in control ("photometric correction") using a "Linear Regression Equation" is also provided.

10 The present invention provides a method (A) for a Calibration Algorithm Transfer comprising:

- (i) obtaining a first set of absorbance measurements of a set of calibrators on a First Apparatus that is in control at wavelengths from a first wavelength calibration table;
- 15 (ii) establishing a second wavelength calibration table on a second apparatus, the first and the second wavelength calibration table may be the same or different, and obtaining a second set of absorbance measurements of the set of calibrators on the Second Apparatus, at wavelengths from the second wavelength
20 calibration table;
- (iii) determining a first interpolated absorbance for the first absorbance measurements for at least one wavelength of a Standard Set of Wavelengths, and determining a second
25 interpolated absorbance for the second absorbance measurements for the at least one wavelength of the Standard Set of Wavelengths,
- (iv) deriving a First Linear Regression Equation for each of the at least one wavelength of the Standard Set of Wavelengths using the first and the second interpolated absorbance
30 measurement;

- (v) incorporating the First Linear Regression Equation and at least one Primary Calibration Algorithm onto the Second Apparatus.

In a further aspect of the invention there is provided a method (B) of
5 determining the concentration of an analyte in a sample in a second
apparatus comprising:

- (a) performing a Calibration Algorithm Transfer, as defined above;
- (b) measuring an absorbance of the sample on the second
apparatus, and determining a sample interpolated absorbance
10 for at least one wavelength of the Standard Set of wavelengths;
- (c) adjusting the interpolated absorbance with the First Linear
Regression Equation to obtain an Adjusted Interpolated
Absorbance; and
- (d) calculating a concentration for the analyte by applying at least
15 one Primary Calibration Algorithm for the analyte to the
Adjusted Interpolated Absorbance.

In yet a further aspect there is provided a method (C) to derive a
standard set of wavelength from a wavelength calibration table wherein the
20 wavelength calibration table for the first or second apparatus is obtained by:

- (i) projecting a first electromagnetic radiation of known wavelength,
onto a first pixel of a first linear diode array of the first apparatus or
a second linear diode array of the second apparatus;
- (ii) using a second electromagnetic radiation of known wavelength,
25 the second electromagnetic radiation having a different
wavelength than the first electromagnetic radiation, projecting the
second electromagnetic radiation onto a second pixel of the first or
second linear diode array;
- (iii) identifying the first and second pixels within the first or second
30 linear diode array;

- (iv) calculating a pixel dispersion for the first and second linear diode array; and
- (v) assigning a wavelength to each pixel within the first and second linear diode array to produce the wavelength calibration table using the pixel dispersion and either the first electromagnetic radiation of known wavelength, and the first pixel, or the second electromagnetic radiation of known wavelength and the second pixel.

10 The wavelength calibration table for the first apparatus may also be obtained by:

- (a) projecting a known wavelength of electromagnetic radiation, onto a pixel in a linear diode array of the first apparatus;
- (b) identifying the pixel number of the pixel;
- 15 (c) assigning a wavelength to each pixel within the linear diode array to produce the wavelength calibration table using a predetermined pixel dispersion, the known wavelength of electromagnetic radiation, and the pixel number.

20 The wavelength calibration table of a second apparatus may be similarly obtained by projecting the electromagnetic radiation of known wavelength onto a pixel of the linear diode array of the second apparatus having the same pixel number as that of the first apparatus. Alternatively, the electromagnetic radiation may be projected on a pixel having a different pixel
25 number.

In another aspect of the invention, the standard set of wavelengths can be obtained by:

- (A) establishing a set of wavelengths common to the wavelength calibration table of both the first and second apparatus; and
- 30

- (B) selecting a range of wavelengths of the set of wavelengths, the range of wavelengths having wavelengths belonging to the standard set of wavelengths.

5 In yet a further aspect of the invention photometric correction is provided by the following equation:

$$AIA = (\text{interpolated absorbance} - y - \text{intercept}) / \text{slope};$$

10 Wherein, "AIA" is the Adjusted Interpolated Absorbance, "interpolated absorbance" is as determined in the step of measuring, see step (b), Method (B) as described above, and "y-intercept" and "slope" are obtained from the first linear regression equation, where the First Linear Regression Equation is derived from a plot of interpolated absorbance measurements, the first
15 interpolated absorbance measurements on an X-axis, and the second interpolated absorbance measurements on a Y-axis, the First linear regression equation having a y-intercept and a slope.

In a further aspect of the invention, a Second Apparatus that was
20 calibrated by "Calibration Algorithm Transfer" but is no longer in control, can be "Recalibrated" using a Set of Calibrators that was assigned absorbances or absorbance values by the First Apparatus, which was known to be in control. The present invention also provides a method for Recalibration of the First Apparatus in the same way as any Second Apparatus.

25

In this aspect of the invention, the Calibrators are measured in the apparatus which is being Recalibrated and absorbances recorded using the same Standard Set of Wavelengths as used in the First Apparatus. New, or
"Second Linear Regression Equations" are then developed for each relevant
30 wavelength with the absorbance measurements from this Set of Calibrators

versus the absorbance measurements assigned by the First Apparatus to the lot or batch of Calibrators. Each generated Second Linear Regression Equation having an intercept and slope, is then stored in the apparatus being Recalibrated. Interpolated absorbance measurements of actual samples in
5 the Recalibrated apparatus are then adjusted using the Second Linear Regression Equation(s). The term "Second Linear Regression" applies to Recalibration (and also calibration as described below) of an apparatus and the term "First Linear Regression" applies to Calibration Algorithm Transfer from a First Apparatus to a Second Apparatus.

10

Thus the invention also provides a method (D) for Recalibrating an apparatus that is no longer in control comprising:

(i) obtaining absorbance measurements of a set of calibrators on the apparatus, the set of calibrators having assigned absorbance values, the
15 apparatus comprising a Primary Calibration Algorithm;

(ii) determining interpolated absorbance values for the absorbance measurements for at least one wavelength of a Standard Set of Wavelengths;

(iii) establishing a Second Linear Regression Equation in the
20 apparatus, using the interpolated absorbance values and the assigned absorbance values; and

(iv) incorporating the Second Linear Regression Equation on the apparatus to produce a recalibrated apparatus.

25 The concentration of an analyte in a sample may also be obtained (Method E) in a Recalibrated apparatus by;

(a) recalibrating the apparatus

(b) measuring an absorbance measurement of the sample;

(c) deriving an interpolated absorbance for the absorbance measurement for at least one wavelength of the Standard Set of Wavelengths in the recalibrated apparatus;

(d) adjusting the interpolated absorbance measurement with the
5 Second Linear Regression Equation to obtain an Adjusted Interpolated Absorbance; and

(e) calculating a concentration for the analyte by applying the Primary Calibration Algorithm for the analyte to the Adjusted Interpolated Absorbance.

10

Photometric correction may also be performed on recalibrated apparatuses using the following equation:

$$\text{AIA} = (\text{interpolated absorbance} - \text{y-intercept}) / \text{slope};$$

15

wherein, "AIA" is Adjusted Interpolated Absorbance, "interpolated absorbance" is as determined in the step of deriving (step (c)), Method (E) as described above, and "y-intercept" and "slope" are obtained from the Second Linear Regression Equation, where the Second Linear Regression Equation
20 is derived from a plot of electronically stored assigned absorbance measurements on an X-axis, and the interpolated absorbance measurements obtained on the recalibrated apparatus on a Y-axis, the Second linear regression equation having a y-intercept and a slope.

25 In a further aspect of the invention there is provided a method (F) for calibrating an apparatus lacking a primary calibration algorithm the comprising:

(i) obtaining absorbance measurements of a Set of Calibrators on the apparatus, the apparatus lacking a primary calibration algorithm, and the
30 set of calibrators having assigned absorbance values,

(ii) determining interpolated absorbance values for the absorbance measurements for at least one wavelength of a Standard Set of Wavelengths;

(iii) establishing a Second Linear Regression Equation in the apparatus, using the interpolated absorbance measurements and the assigned absorbance values; and

(iv) incorporating the Second Linear Regression Equation, and at least one Primary Calibration Algorithm on the apparatus, to produce a calibrated apparatus.

10

A calibrated apparatus may be used to determine (Method G) the concentration of an analyte by:

(a) calibrating the apparatus according to the method described above;

15 (b) measuring an absorbance value of the sample;

(c) deriving an interpolated absorbance from the absorbance value for at least one wavelength of the Standard Set of Wavelengths in the calibrated apparatus;

(d) adjusting the interpolated absorbance measurement with the Second Linear Regression Equation to obtain an Adjusted Interpolated Absorbance; and

(e) calculating a concentration for the analyte by applying the Primary Calibration Algorithm for the analyte to the Adjusted Interpolated Absorbance, and wherein in the step of adjusting the interpolated absorbance is obtained using the following equation:

25

$$AIA = (\text{interpolated absorbance} - y\text{-intercept}) / \text{slope};$$

wherein, "AIA" is Adjusted Interpolated Absorbance, "interpolated absorbance" is as determined in the step of deriving (step (c)) of method (G)

30

as described above and "y-intercept" and "slope" are obtained from the Second Linear Regression Equation, where the Second Linear Regression Equation is derived from a plot of electronically stored assigned absorbance measurements on an X-axis, and the interpolated absorbance measurements
5 obtained on the calibrated apparatus on a Y-axis, the Second linear regression equation having a y-intercept and a slope.

The inventor has also found that the process of Calibration Algorithm Transfer and subsequent determination of analyte concentration can be
10 accomplished by using an order of derivative of the absorbance in the Primary Calibration Algorithm, where absorbance correction or "Photometric Correction" may not be necessary, provided that the order of derivative of absorbance used in the Primary Calibration Algorithm at the selected wavelength(s) does not contain significant inter-apparatus variability as may
15 be seen in the absorbances at the same wavelength(s). Absorbance variability between apparatus can be minimized in certain "Sections of the Absorbance Spectra," by using an order of derivative of the absorbance.

Thus in another aspect of the invention there is provided a method (H)
20 of determining the concentration of an Analyte in a Sample in a second apparatus comprising:

- (i) incorporating at least one primary calibration algorithm that uses an order of derivative of absorbance obtained for at least one of a standard set of wavelengths, on the second apparatus;
- 25 (ii) measuring absorbance values of the sample at three or more wavelengths from a wavelength calibration table on the second apparatus;
- (iii) determining interpolated absorbance values from the absorbance values for wavelengths from a standard set of wavelengths;
- (iv) obtaining a derivative of the interpolated absorbance values,
30 using the order of derivative; and

(v) calculating a concentration of the Analyte in the sample, by applying the Primary Calibration Algorithm to the derivative.

In another aspect of the invention there is provided a method wherein
5 absorbance values may be assigned to a second batch of calibrators using an apparatus that is in control. In particular a "Second apparatus" may be used wherein adjusted interpolated absorbances are assigned to the calibrators of the second batch.

10 According to another aspect of the invention the Primary Calibration Algorithms and the absorbance measurements of the calibrators made on the First Apparatus can be electronically stored. Thus in a further aspect of the invention there is provided a medium for storing instructions adapted to be executed by a processor to determine analyte concentration within a sample,
15 the instructions comprising

- i) at least one primary calibration algorithm;
- ii) the assigned absorbances of a set of calibrators obtained from a first apparatus; and
- iii) the identity of first apparatus used to obtain the at least one
20 primary calibration algorithm and the assigned absorbances.

In yet a further aspect of the invention there is provided a kit comprising the set of calibrators and the medium with the stored instructions as described above.

25

In another aspect of the invention there is also provided an apparatus for determining analyte concentration of a sample comprising a spectrophotometer, a light source, a power supply, a sample holder, a circuit board, a primary calibration algorithm, and said first linear regression
30 equation.

In a further aspect of the invention there is also provided a system for determining presence of an analyte comprising

- 5 i) means for transmitting electromagnetic radiation of one or more known wavelengths through a sample;
- ii) means for detecting electromagnetic radiation after transmission through the sample;
- iii) means for incorporating a primary calibration algorithm,
- 10 iv) means for storing a wavelength calibration table and a standard set of wavelengths;
- v) means for deriving a first linear regression equation, a second linear regression equation, or both a first and a second linear regression equation;
- vi) means for detecting presence or concentration of an analyte within the
- 15 sample.

The present invention provides a method to provide a simple reliable method of using primary calibration algorithms in a second apparatus that do not require representative samples for which the apparatus was designed.

20 Rather, the standard samples used to calibrate a second apparatus can be any stable samples that produce a range of absorbances at all relevant wavelengths.

This summary of the invention does not necessarily describe all

25 necessary features of the invention but that the invention may also reside in a sub-combination of the described features.

Other features and advantages of the present invention will become apparent from the following detailed description. It should be understood,

30 however, that the detailed description and the specific examples while

indicating preferred embodiments of the invention are given by way of illustration only, since various changes and modifications within the spirit and scope of the invention will become apparent to those skilled in the art from this detailed description.

5

BRIEF DESCRIPTION OF THE DRAWINGS

The invention will now be described in relation to the drawings in which:

Figure 1 is a graphic representation of the absorbance spectra of four
10 different synthetic calibrators, tested on the First Apparatus;

Figure 2 is a graphic representation of the absorbance spectra of four different synthetic calibrators, tested on the Second Apparatus;

Figure 3 is scatter plot of the absorbances of the four calibrators at 500nm, when tested on the First Apparatus (x-axis) and the Second
15 Apparatus (y-axis);

Figure 4 is scatter plot of the absorbances of the four calibrators at 550nm, when tested on the First Apparatus (x-axis) and the Second Apparatus (y-axis);

Figure 5 is scatter plot of the absorbances of the four calibrators at
20 600nm, when tested on the First Apparatus (x-axis) and the Second Apparatus (y-axis);

Figure 6 is scatter plot of the absorbances of the four calibrators at 650nm, when tested on the First Apparatus (x-axis) and the Second Apparatus (y-axis);

25 Figure 7 is scatter plot of the absorbances of the four calibrators at 700nm, when tested on the First Apparatus (x-axis) and the Second Apparatus (y-axis);

Figure 8 is scatter plot of the absorbances of the four calibrators at 750nm, when tested on the First Apparatus (x-axis) and the Second
30 Apparatus (y-axis).

Figure 9 is a graphic representation of the absorbance spectra of the same synthetic fluid tested on 3 different apparatuses (#2, #4, #6) using a standard set of wavelengths.

Figure 10 is a graphic representation of the first derivative of absorbance spectra of the same synthetic fluid tested on 3 different apparatuses (#2, #4, #6) using a standard set of wavelengths; the absorbance spectra are shown in Figure 9.

Figure 11 is a scatter plot of the predicted Hb using the calibration algorithm of Example 2 in g/L.

Figure 12 is a scatter plot of the predicted Hb in g/L Hb using the calibration algorithm of Example 2.

Figure 13 is a scatter plot of the predicted Hb in g/L Hb using the calibration algorithm of Example 2.

Figure 14 is a scatter plot of the predicted Hb in g/L Hb using the calibration algorithm of Example 2.

Figure 15 is a scatter plot of the predicted Hb in g/L Hb using the calibration algorithm of Example 2.

Figure 16 is a graphic representation of the second derivative of absorbance spectra of the same synthetic fluid tested on 3 different apparatuses (#2, #4, #6) using a standard set of wavelengths.

Figure 17 is a graphic representation of the second derivative of absorbance spectra of the same synthetic fluid tested on 3 different apparatuses using a standard set of wavelength, and shifting the second derivative of absorbances for apparatus #6 to the right by 2 pixels.

Figure 18 is a graphic representation of the first derivative of absorbance spectra of the same synthetic fluid tested on 3 different apparatuses (#2, #4, #6) using a standard set of wavelength, and shifting the first derivative of absorbances for apparatus #6 to the right by 2 pixels.

Figure 19 is a graphic representation of the third derivative of absorbance spectra of the same synthetic fluid tested on 3 different apparatuses (#2, #4,

#6) using a standard set of wavelengths. The wavelengths identified as 593nm and 608nm were used in the calibration algorithm of Example 2 for hemoglobin when the first derivative of absorbance was used.

5 **DETAILED DESCRIPTION OF INVENTION**

The following description is of a preferred embodiment by way of example only and without limitation to the combination of features necessary for carrying the invention into effect.

10 The method used to calibrate a first apparatus wherein said apparatus can be used to measure the concentration of at least one analyte is referred to as Primary Calibration. Primary Calibration is a complex process and is described under the title "Primary Calibration." Due to its complexity, performance of primary calibration on every apparatus is not desirable.

15

The present invention provides a simple alternative that allows an apparatus to function as though it was calibrated by the process of primary calibration, whereby the apparatus need not be calibrated in the same way in which the first apparatus was calibrated, that is by conducting a primary
20 calibration. It is preferred that the first and second apparatus are similar. There are several aspects of this invention including:

- i) Calibration Algorithm Transfer
- ii) Recalibration
- iii) Calibration, and
- 25 iv) Calibration of a Second Apparatus Using Derivative of Absorbance

These four, as well as other, aspects of the invention are described in detail below.

Where calibrators are required for the invention, the calibrators are
30 preferably synthetic calibrators, but other calibrators may be used as will be

obvious as the invention is described in details. With respect to synthetic calibrators, part of the process of calibration algorithm transfer teaches one method of assigning absorbances to a batch of calibrators. It is preferred that an apparatus, for example, the first apparatus, that is in control be used.

5

A fifth aspect of the invention that includes use of a second apparatus is as follows:

v) Assigning Absorbances to a Second Batch of Calibrators

Methods for assigning absorbances to a second batch of calibrators are described, wherein the original primary calibration algorithm(s) may be used
10 whereby the process of primary calibration could be avoided.

The first and second apparatus use absorption of electromagnetic radiation to measure analytes in one or more samples. Throughout this
15 application reference is made to measurements through absorption of electromagnetic radiation. While this is preferred, it should be understood that the present invention is not limited to absorption and measurements by other means such as reflectance are also within the scope of the present invention.

20

Technical terms used in the disclosure are defined below for clarification.

By "Actual Wavelength" it is meant a wavelength from the wavelength calibration table.

25

By "Analyte" it is meant a substance being measured in a sample.

By "absorbance", it is meant a measurement calculated from the amount of light reflected or transmitted by a sample, as would be known to one of skill
30 in the art.

By "Assigned Absorbances" or "Assigned Absorbance Measurement" or "Assigned Absorbance Values" it is meant the interpolated absorbances of calibrators whose absorbances are measured on the first apparatus when it is in control, or the adjusted interpolated absorbances of a second batch of calibrators.

By "Actual Absorbance" or "Measured Absorbance" it is meant the absorbance value, or absorbance measurement, or simply absorbance of a sample or calibrator provided by the apparatus at a wavelength from the wavelength calibration table of the apparatus.

By "Adjusted Interpolated Absorbance" it is meant the value of the interpolated absorbance after photometric correction is applied specifically to the interpolated absorbance.

By "Blood Bag Tubing" it is meant the tubing connecting a first plastic bag that contains whole blood and a second plastic bag that may contain plasma obtained from said first bag. The tubing and bags may be made from transparent or translucent flexible plastic.

By "Calibration Set", a "Set of Calibrators", or "Calibrators" it is meant two or more calibrator samples. Any type of Calibrators suitable for producing one or more first linear regression equations based on interpolated absorbances obtained using a first apparatus that is in control and interpolated absorbances obtained using a Second apparatus, may be used.

By "Calibration" it is implied the process of incorporating at least one primary calibration algorithm in an apparatus that does not have at least one valid primary calibration algorithm incorporated, plus the process of

establishing and incorporating one or more second linear regression equations in said apparatus. Calibration does not require the first apparatus. Typically, calibration of an apparatus is required when said apparatus does not have at least one valid relevant primary calibration algorithm
5 incorporated.

By "Calibration Algorithm Transfer" or "Transferring a Calibration Algorithm" it is meant the process of incorporating at least one primary calibration algorithm and one or more first linear regression equation within a
10 second apparatus. A first linear regression equation is necessary for performance of photometric correction. Calibration algorithm transfer requires the first apparatus, a second apparatus, and a set of calibrators.

By "Derivative of Absorbance" it is meant an order of derivative of the
15 absorbance spectrum. For example, the first order derivative of absorbance at a particular wavelength is the slope of the absorbance spectrum at the wavelength; the second order derivative of absorbance at a particular wavelength is the slope of the first derivative absorbance spectrum at said wavelength. Methods of calculating a derivative of absorbance at a particular
20 wavelength are well known by those skilled in the art. The first derivative of absorbance at a particular wavelength may comprise the difference in absorbances at the two wavelengths that encompass said wavelength. Other methods of calculating derivative of absorbance may use the absorbances at several different wavelengths, where smoothing is an integral part of the
25 derivative process. It should be understood that with a greater degree of smoothing, there is also a greater loss of signal details in the absorbance spectrum or derivative of absorbance spectrum. The minimum number of wavelengths that may be used to calculate a derivative of absorbance is two wavelengths; the order of that derivative is first order.

By "Discrete wavelength" it is meant a single wavelength.

By "First Absorbance" it is meant the absorbance measured on the first apparatus.

5

By "First Apparatus" it is meant an apparatus used to develop the at least one primary calibration algorithm.

"A First Linear Regression Equation" is an equation of the form " $y = mx + c$ " obtained from the absorbances obtained from a set of calibrators on both
10 the first apparatus, and a second apparatus during the process of calibration algorithm transfer, where " m " is the slope and " c " is the y-intercept.

By "Full spectrum" it is meant continuous sections of the absorbance
15 spectrum or derivative of absorbance spectrum, up to and including a complete spectrum. For example, which is not to be considered limiting in any manner, the absorbance of primary calibrators over a full spectrum may be used to determine a primary calibration algorithm, for example, using Partial Least Squares ("PLS") analysis or Principal Component Analysis
20 (PCA). It should be understood that sections of a full spectrum may also be used for PLS analysis or PCA.

By "Interpolated Absorbance," it is meant the absorbance value for a specific wavelength of a standard set of wavelengths. If the specific
25 wavelength of said standard set of wavelengths is the same as the wavelength already assigned to a pixel of a linear diode array, i.e., a wavelength of the wavelength calibration table, then the interpolated absorbance is the same as the measured absorbance. However, if the specific wavelength is different from the wavelength assigned to a pixel, then
30 the value of interpolated absorbance, at the specific wavelength, is estimated

from the measured absorbance values of at least two adjacent pixels whose wavelengths encompass one or more standard wavelengths. It should be noted that the term "Interpolated Absorbance" is applied when the wavelength calibration table is the same as the standard set of wavelengths, and also when the wavelength calibration table is different from the standard set of wavelengths, i.e., whether or not there was a need for interpolation.

An apparatus is said to be "In Control" when that apparatus produces absorbance measurements that are within a predetermined acceptable range for one or more given wavelengths, or predicted analyte values that are within a predetermined acceptable range for one or more analytes.

By "Interferent" it is meant an analyte whose presence in a sample, particularly a serum or plasma sample, interferes with the determination of the presence and/or quantification of another analyte(s).

By "Mapping" it is meant the process of associating an interpolated absorbance value with a standard wavelength.

By "Optimizing Primary Calibration Wavelengths" it is meant a process of choosing the Primary Calibration Wavelength(s) that exhibits measurable changes in the derivative absorbance as a function of analyte concentration (measurable analyte signal) and which also exhibits a low variability, that is to say a variability that is below a predetermined value in the derivative of absorbance between apparatuses for a given analyte concentration, for each of the optimized calibration wavelengths. Examples of low variability values, include values that translate into less than 20%, preferably less than 10%, and more preferably less than 5% coefficient of variation (CV) of predicted analyte concentration, where:

$$\%CV = ([\text{Standard Deviation} / \text{Mean}] \times 100\%).$$

Optimization of primary calibration wavelengths may be accomplished by one of two methods that comprises:

- 5 1) Plot the absorbance spectra vs the standard set of wavelengths of a sample tested on two or more apparatus, as exemplified in Figure 9; plot the derivative of absorbance of said sample tested on two or more apparatus, as exemplified in Figure 10, Figure 16 and Figure 19 for the first, second and third derivative or absorbance respectively. Visually
10 select one or more wavelength of a standard set of wavelengths that exhibit a low variability in the derivative of absorbance and also with measurable analyte signal, preferably with least variability in the derivative of absorbance and also with measurable analyte signal.
- 15 2) When multiple linear regression is used to develop the primary calibration algorithm, a method used to select wavelengths that show minimum variability in the absorbance, or derivative of absorbance, is to test the primary calibration set on two or more apparatus, preferably apparatus that exhibit the expected limit of absorbance and wavelength variability, and use the data from all the apparatus in the
20 primary calibration data set. An example of a software tool used for preparing primary calibration algorithms is StatView™ that is used to create calibration algorithms by a process called step-wise multiple linear regression. In step-wise multiple linear regression, absorbance or derivative of absorbance measurements for all the wavelengths may
25 be presented to the StatView™ program, and only the wavelengths at which the criteria of changes and low variability in derivative of absorbance are met would be selected for the primary calibration algorithm based upon how well a primary calibration algorithm for a particular analyte predicts the analyte concentrations of the samples of
30 the primary calibration set. Further, wavelengths which contribute

least to the calibration algorithm may be eliminated until at least the desired number of wavelengths are left. After the wavelength(s) is/are selected, preferably only the data generated from a first apparatus should be used to develop the final primary calibration algorithm, using the predetermined wavelength(s). Preferably more than the derived number of wavelengths are submitted to StatView™ software, with data only from the first apparatus, and further elimination of wavelengths not meeting the criteria should be performed until the desired number of wavelengths remain.

10

By "Primary Calibration Algorithm" it is meant a process used to develop a primary calibration algorithm for a first apparatus for an analyte. The sample set used for calibration is relatively large, and the samples are natural or very close to natural samples. The primary calibration set should include all the variability expected in a sample, in order to develop robust calibration algorithm(s).

15

By a "Primary Calibration Set" it is meant the samples used for primary calibration.

20

By "Primary Calibration Wavelength(s)" it is meant the wavelength(s) used in a primary calibration algorithm.

By "Pixel dispersion" it is meant, the wavelengths encompassed by two adjacent pixels of a linear diode array, usually measured in nanometers (nm) per pixel. For example, if two lasers of 600nm and 900nm are used for wavelength calibration, and they are projected on pixel 20 and pixel 220 respectively, that means 300 nm (i.e., 900 - 600nm) are encompassed by 200 pixels (i.e., 220 - 20 pixels). Therefore the Pixel dispersion is calculated to be

25

1.5nm per pixel (i.e., 300nm divided by 200 pixels). The assigned Pixeldispersion may be predetermined by the steps comprising:

- (i) Projecting a first electromagnetic radiation of known wavelength, onto a first pixel of a linear diode array preferably towards one end of the linear diode array;
- (ii) Using a second electromagnetic radiation of a known wavelength but different from the first electromagnetic radiation, projecting the second electromagnetic radiation onto a second pixel that is preferably towards the opposite end of the linear diode array;
- (iii) Identifying said first and second pixels within the linear diode array;
- (iv) Calculate the Pixeldispersion in nanometers per pixel or any wavelength units per pixel as described above; and

Use the Pixeldispersion value calculated in step (iv) as the predetermined Pixeldispersion, or repeating steps (i) to (iv) on one or more apparatus with linear diode arrays, preferably the linear diode arrays of the apparatus are similar, and use the average pixeldispersion as the predetermined Pixeldispersion.

By "Photometric Correction" or "Absorbance Adjustment" it is meant an adjustment made to an absorbance of a sample tested on one apparatus, to make it appear as if the sample was tested on another apparatus. The amount of photometric correction is determined by the slope and y-intercept of the first and second linear regression equations. The resulting absorbance after photometric correction is referred to as Adjusted Absorbance or Corrected Absorbance.

By "Relevant Primary Calibration Algorithm" it is meant the primary calibration algorithm for a particular analyte, particularly with respect to the one installed in an apparatus. A primary calibration algorithm is required for each analyte, regardless of the wavelength or wavelengths used in the

primary calibration algorithm(s), and one or more primary calibration algorithms may be installed in an apparatus.

By "Recalibration" it is meant a process of establishing and incorporating
5 one or more Second Linear Regression Equations in an apparatus that already has at least one valid primary calibration algorithm incorporated. Typically, recalibration of an apparatus is required when said apparatus is not In Control.

10 By "Sample" or "Samples" it is meant a biological or non-biological fluids, or a solid exhibiting one or more properties that may be measured spectrophotometrically. A sample typically comprises one or more analytes. A sample is characterized in that the energy of one or more wavelengths can be either transmitted, absorbed, scattered, reflected, or a combination
15 thereof, by an analyte within the sample. Examples of a sample include, but are not limited to, a calibrator, serum or plasma, or a pharmaceutical tablet.

By "Smoothing" a curve, for example an absorbance spectrum, it is meant applying a mathematical function to the digital data to produce a "continuous
20 spectrum" and thereby reduce the "noise" in the spectrum. Various degrees of smoothing may be applied to a curve. The loss of analyte signal may be a price paid for smoothing.

By "Second Apparatus" it is meant an apparatus that is allowed to function
25 like a first apparatus, whereby the second apparatus need not be calibrated in the same way in which the first apparatus was calibrated, i.e., by conducting a primary calibration.

"A Second Linear Regression Equation" is an equation of the form " $y = mx$
30 $+ c$ " obtained from the interpolated absorbances obtained from a set of

calibrators on both the first apparatus and any other apparatus during the process of "Recalibration," or "Calibration" where "m" is the slope and "c" is the y-intercept.

- 5 By "Second Absorbance" it is meant the absorbance measured on the second apparatus.

By a "Section of the Absorbance Spectra" it is meant absorbances at more than one wavelength where the wavelengths are in succession.

10

- By a "Standard Set of Wavelengths" it is meant a set of wavelengths used by all apparatus in conjunction with the apparatus-specific wavelength calibration table, used to generate interpolated absorbances from the measured or actual absorbances. The actual absorbances of a sample
- 15 tested on an apparatus are measured at wavelengths from the wavelength calibration table, and said actual absorbances are interpolated and mapped onto the standard set of wavelengths. The primary calibration algorithm(s) is/are preferably applied to the mapped absorbances, but may be applied to the actual absorbances, particularly when the wavelength calibration table
- 20 and the standard set of wavelengths are the same. Without wishing to be limiting in any manner, an example of a standard set of wavelengths includes 500 to 1100nm, in increments of 2nm. However, other wavelength ranges and increments may be used as required, and as would be known by one of skill in the art. The range of the standard set of wavelengths may be derived
- 25 from the wavelength calibration table, and the increment may be obtained by trial and error. The standard set of wavelengths may also be obtained by establishing a set of wavelengths common to the wavelength calibration tables of both first and second apparatus.

By a "Standard Wavelength" it is meant a wavelength from the standard set of wavelengths.

By a "Validation Set" it is meant the samples used for testing the prediction accuracy of a primary calibration algorithm, and may also be used for testing the robustness of said primary calibration algorithm.

By "Vessel" it is meant any transparent or translucent container capable of holding a sample, preferably fluid, to enable measurement of absorbance, reflectance, or both absorbance and reflectance of radiation in the sample. Examples of vessels include, but are not limited to, pipette tips, tubing, cuvettes, labeled test tubes, unlabeled test tubes, blood bag tubing, any transparent sample container, and any translucent sample container. In the case of a solid sample, a sample holder may be required instead of a Vessel.

15

By "Valid Primary Calibration Algorithm" it is meant that the apparatus is in control, particularly for the analyte predicted by said primary calibration algorithm.

By "Wavelength Calibration" it is meant the calibration of a Linear Diode Array detector of a spectrophotometer, wherein wavelengths are assigned to each pixel in said Linear Diode Array.

By "Wavelength Calibration Table" it is meant a table that provides the actual wavelength corresponding to or assigned to each pixel, which is a result of the wavelength Calibration.

The present invention provides a method to calibrate a second apparatus using a Primary Calibration algorithm(s) developed for a First Apparatus. Primary calibration may be developed on the first apparatus using known

30

techniques. The present invention also provides a method for Recalibrating a second apparatus, and if required, recalibration of the First Apparatus.

It is preferred that the Primary Calibration Algorithm(s) and the assigned
5 absorbance measurements of calibrators are stored on a suitable medium, for example, electronically stored, permitting ready transfer to the Second Apparatus.

It is also preferred that the samples used for the Primary Calibration, i.e.,
10 the primary calibration set, and the calibrators, are placed within vessel having optical properties substantially similar. Examples of vessel include, but are not limited to, a pipette tip, a labeled test tube, an unlabeled test tube, a blood bag tubing, any transparent container, or any translucent container. The sample can be liquid for example, a biological fluid, including but not
15 limited to bodily fluids, serum or plasma, or non-biological, for example but not limited to synthetic calibrators or a beverage, or solid, for example, which is not to be considered limiting in any manner, cheese or a medicament, for example a pharmaceutical tablet. Furthermore, any type of calibrator suitable for producing a set of first linear regression equations based on assigned
20 absorbances obtained from a First apparatus and interpolated absorbances from a Second apparatus, may be used. For example, a set of calibrators may comprise translucent substances such as fiberglass which may vary in thickness, color, or finish, depending on the type of analyte to be measured. The calibrators may also comprise solid particles in liquid suspension, e.g.,
25 polystyrene beads in a buffer of similar density as the polystyrene beads. Calibrators may also comprise fat emulsion, for example fat particles in suspension. Calibrators may also comprise colored glass of varying thickness and finish. In any event, the calibrators must be capable of producing a suitable set of first linear regression equations for relevant
30 wavelengths, obtained from a First Apparatus and a Second Apparatus.

The apparatus are used to measure concentrations of analytes using electromagnetic radiation, most preferably using near infrared radiation and the adjacent visible radiation spectrum, for example, but not limited to, any
5 suitable wavelength from about 300nm to about 2500nm, preferably from about 500nm to about 1100nm, and more preferably from about 500nm to about 800nm.

A standard set of wavelengths may be used as required for standardizing
10 the absorbance data to be used for obtaining a primary calibration algorithm. The actual set of wavelengths provided by an apparatus depends on the manner in which the wavelength calibration is performed. Further, a standard set of wavelengths and the method used to generate them depends on the accuracy requirement for analyte prediction. In order for a second apparatus
15 to use the primary calibration algorithm(s) developed on a first apparatus, it is preferred that each apparatus should provide the same set of actual wavelengths. An example of a way whereby one or more apparatus can be made to operate as if they have the same wavelengths, which is not to be considered limiting in any manner, is provided under the title "Wavelength
20 Calibration."

A linear wavelength table comprising a known set of wavelengths that fall within the required range and with equal increments, may be used as a standard set of wavelengths. For example, the range may be 500nm to
25 1100nm, in increments of 2nm. The absorbance values measured at the actual wavelengths assigned to each pixel of the apparatus, i.e., the wavelength calibration table, may be interpolated and mapped onto said standard set of wavelengths, thereby allowing different apparatus the appearance of having been provided the same set of actual wavelengths.
30 The number of wavelengths in the standard set of wavelengths do not have

to be equal to the number of pixels in the diode array. For example, the number of pixels may be more or less than the number of wavelengths, but it is preferred that the number of wavelengths should approximate the number of pixels. In a preferred embodiment, similar linear diode arrays with the
5 same number of pixels should be used.

The following is a preferred embodiment of the apparatus used in the present invention:

APPARATUS:

10 The preferred embodiment of the apparatus is described both in general terms and specific requirements. The present invention provides an apparatus that may be used for measuring the concentration of an analyte in a sample. An example, which is not to be considered limiting, of such an apparatus and the primary calibration algorithms is given in
15 WO9838961. It is preferred that apparatus have similar components. The major components of an apparatus may comprise:

1. A spectrophotometer comprising:

A diffraction grating, focussing lenses, slits, and a linear diode array detector (sometimes referred to as simply a linear diode array). It is
20 preferred that spectrophotometers have similar components, including the number of pixels in the linear diode array. It should be understood that other arrays, for example but not limited to CCD (charged coupled detector) arrays, may also be used as described herein. The spectrophotometer can operate in a single or dual beam configuration.
25 If the spectrophotometer is a dual beam spectrophotometer, one of the beams acts as a reference beam and the other is the sample beam; two shutters are required to facilitate the sample and reference measurements.

2. A light source

30 3. A power supply

4. A sample holder with light coming from the lamp through an optical fiber, and light transmitted through the sample to the spectrophotometer via a second optical fiber.
5. A circuit board that includes an amplifier and an analog to digital converter, is required to interface the linear diode array detector and a microprocessor. The Primary Calibration Algorithm and other information like the wavelength calibration table may be imbedded in an EPROM.
6. Software that may comprise features for: utilizing the calibration package; interpolating absorbances; mapping absorbances to a standard set of wavelengths; smoothing; creating derivatives of absorbances; calculating analyte concentrations.

Differences in absorbances for the same sample occur between apparatus for a number of reasons including:

1. The corresponding pixels in similar linear diode array detectors respond differently to the same amount of light.
2. The incremental wavelength per pixel or pixel dispersion may vary between any two similar linear diode array detectors.
3. The physical distances between pixels in the linear diode array are not always constant within a linear diode array, and also between similar linear diode arrays.
4. Spectrophotometers contain wavelength inaccuracies, depending on the method used for wavelength calibration. One illustration is provided in Example 6.
5. Variability in other components of the spectrophotometers, e.g., the diffraction grating, filters, and slits.

The present application is directed to providing a method for calibrating an apparatus for an analyte involving the steps of:

A) WAVELENGTH CALIBRATION: A process of wavelength calibration provides a wavelength calibration table that gives the wavelengths assigned to each pixel. The linear diode array detector comprises of
5 pixels, for example 128 or 256 pixels assembled in a linear array. In order that different apparatus would appear to have the same wavelength calibration table, the absorbances of a sample may be interpolated and mapped unto a standard set of wavelengths. It is preferred that the same standard set of wavelengths be used with each apparatus. When the
10 disclosure refers to "measured at a standard set of wavelengths," it should be understood that "interpolated absorbances mapped to a standard set of wavelengths" is implied, unless all apparatus have the same wavelength calibration table, in which case the step of interpolation is automatically eliminated.

15

B) PRIMARY CALIBRATION: Primary calibration is a process used to develop a primary calibration algorithm for a first apparatus for an analyte. The sample set used for primary calibration is relatively large, and the samples are natural or very close to natural samples and should include
20 all the variability expected in a sample. For each unique analyte a unique primary calibration algorithm is required, therefore, an apparatus may comprise more than one primary calibration algorithm depending on the number of analytes to be measured.

C) VALIDATION: After the calibration algorithm is developed for an
25 analyte, a second sample set is used to validate said calibration algorithm. If enough variability is not included in the primary calibration set, the analyte predictions for the validation set may correlate poorly with the actual analyte concentrations. Proper validation is a test of robustness of
30 the primary calibration algorithm.

D) CALIBRATION ALGORITHM TRANSFER: This section describes the process of calibration algorithm transfer. It is the process of incorporating at least one primary calibration algorithm and one or more first linear regression equation within a second apparatus. A first linear regression equation is necessary for performance of photometric correction. Calibration algorithm transfer requires the first apparatus, a second apparatus, and a set of calibrators.

E) RECALIBRATION: This section describes the sequel to calibration algorithm transfer to a second apparatus, requiring at least two calibrators with absorbance values assigned to them by the apparatus used to develop the primary calibration algorithm, while the apparatus was in control. The first apparatus is not required, unless it is the apparatus being recalibrated, and the primary calibration algorithm will have already been incorporated in the second apparatus. By recalibration it is meant a process of establishing and incorporating one or more Second Linear Regression Equations in an apparatus that already has at least one valid primary calibration algorithm incorporated. Typically, recalibration of an apparatus is required when said apparatus is not in control.

F) CALIBRATION: This section describes the process of incorporating at least one primary calibration algorithm in an apparatus that does not have a at least one valid primary calibration algorithm incorporated, plus the process of establishing and incorporating one or more second linear regression equations in said apparatus. Calibration does not require the first apparatus, unless it is the apparatus being calibrated. Typically, calibration of an apparatus is required when said apparatus does not have at least one valid relevant primary calibration incorporated.

G) CALIBRATION OF A SECOND APPARATUS USING DERIVATIVE OF ABSORBANCE:

Determination of analyte concentration in a sample in a second apparatus may be accomplished by using an order of derivative of absorbance in the primary calibration algorithm, and photometric correction may or may not be necessary, depending on the required accuracy of the predicted analyte concentration.

H) SYNTHETIC CALIBRATION FLUIDS:

Although calibrators can take on several different forms, provided that the set cover a range of absorbances at each relevant wavelength, whereby a linear regression equation between the absorbances from two apparatus can be established for each relevant wavelength. In a preferred embodiment, ready-to-use stable synthetic calibration fluids are used.

I) ASSIGNING ABSORBANCES TO A SECOND BATCH OF CALIBRATORS:

This section describes the use of the first apparatus and a second apparatus to assign absorbance values to a new lot of synthetic calibrators.

J) CALIBRATION PACKAGE:

When synthetic calibration fluids are used for recalibration or calibration of an apparatus, it is preferred that electronic information accompany the synthetic calibration fluids in the form of a package. The calibration package may comprise the following essential components: (i) At least one primary calibration algorithm; (ii) Calibrators; and (iii) Assigned absorbances for calibrators. This section describes the contents of this package in a preferred embodiment.

These steps are described in more details below:

A) WAVELENGTH CALIBRATION

A laser of known wavelength or electromagnetic radiation transmitted through a band-pass filter of known wavelength, is projected onto any pixel in a linear diode array. It should be understood that the electromagnetic radiation should not be restricted to a laser or a band-pass filter, and other sources of monochromatic electromagnetic radiation may be used. It should also be understood that the electromagnetic radiation could impinge upon more than one pixel, and that the relative position of peak intensity of the electromagnetic radiation may be determined mathematically by processes known to those skilled in the art. Further, the peak intensity may be positioned between any two pixels. The targeted pixel is preferably towards one end of the spectrum. A second laser of known wavelength or electromagnetic radiation transmitted through a second band-pass filter of known wavelength that is preferably projected towards the other end of the spectrum may be used and the pixel on which the beam is projected onto is identified. Since the number of pixels is known, one can determine the pixel dispersion. With the two known wavelengths and their corresponding pixels, and the pixel dispersion, one can generate a wavelength calibration table i.e., a table providing the discrete wavelength that is assigned to each pixel in the linear diode array. The absorbances at the wavelengths from the wavelength calibration table from one or more apparatus, can subsequently be interpolated and mapped onto a standard set of wavelengths. The absorbances at the two actual wavelengths that are on either side of the standard wavelength may be interpolated to produce an absorbance at a standard wavelength. This process may be repeated for each standard wavelength. This is, the preferred method for making the wavelengths provided by different apparatus, appear similar. Photometric accuracy depends in part on wavelength accuracy, and the prediction accuracy for an analyte concentration depends upon the photometric accuracy of the apparatus. In this respect, a qualitative method for an analyte where a yes/no

answer is all that is desired does not require the same level of wavelength accuracy as a quantitative method for the same analyte.

In this method of wavelength calibration, the first wavelength does not
5 have to be projected upon the same pixel in the linear diode array of each
apparatus, since the absorbances could be interpolated and mapped unto a
standard set of wavelengths. The wavelength of a second laser or second
band-pass filter is preferably chosen so that the beam of electromagnetic
radiation is projected towards the other end of the linear diode array. It is
10 preferred that the laser or band-pass filter be selected so that the beam of
electromagnetic radiation is not projected too close to the end pixels in the
linear diode array, if the resulting absorbances at the end pixels are noisy. It
is also preferred that a bandpass filter is a narrow bandpass filter.

15 A second method to generate a wavelength calibration table is to project
the first beam onto the same pixel of each linear diode array. When this
method is used to generate a wavelength calibration table, the
pixeldispersion is predetermined using two beams of different wavelengths,
as described above. The pixeldispersion may be determined from a single
20 spectrophotometer, but preferably the average value should be obtained from
more than one like spectrophotometer. When the same pixeldispersion is
used by each apparatus and the first beam is projected onto the same pixel
number within each like linear diode array, the wavelength calibration table
for each apparatus would be the same, and hence said wavelength
25 calibration table may be used as the standard set of wavelengths.
Consequently interpolation and mapping of absorbances to a standard set of
wavelengths would automatically be eliminated. A second beam may be
used to validate wavelength accuracy.

A third method to generate a wavelength calibration table is like the second method except that the first beam may be projected onto any pixel of the linear diode array. When the pixel number that the first beam is projected onto, is different in different apparatus, the pixel numbers assigned to a specific wavelength in the wavelength calibration table of the different apparatuses will differ. In this case, software may be used to produce a standard set of wavelengths as follows:

- (i) Establish a set of wavelengths common to the wavelength calibration table of the different apparatus.
- (ii) Select a range of wavelengths of the standard set of wavelengths, said range of wavelengths having wavelengths belonging to the standard set of wavelength.

It should be understood that the wavelength calibration table obtained from different apparatus as described in above third method may be such that a pixel number from different apparatus may not be assigned the same wavelength. It should also be understood that the first pixel may be an approximation to a pixel number and also said first pixels from different apparatus may be approximated to be the same pixel, and that the approximations tolerated depends on the prediction accuracy required for the primary calibration algorithms. In other words, the identification of the first pixel may be incorrect. An incorrect identification can be tolerated provided that the incorrectly identified pixel is within less than or equal to about $\pm N$ pixel, where N is the number of pixels that encompass a range of wavelength. Fore example, if the pixel dispersion is 2nm and if the tolerated error is ± 10 nm, then the incorrectly identified pixel must be no more than 5 pixels away on either side of the actual pixel on which the beam impinged. Different levels of error may be tolerated typically, but not limited to ± 2 nm to ± 20 nm and more preferably from ± 2 nm to ± 10 nm. Selection of a wavelength

calibration method depends on the required prediction accuracy of the primary calibration algorithms.

For the purpose of calibration, the actual wavelengths may be irrelevant, if the number of wavelengths in the standard set of wavelength is equal to the number of pixels of the linear diode array, and the pixels in each linear diode array are equivalent in terms of wavelength. The identification of the actual wavelengths associated with each pixel is not necessary to generate calibration algorithms. In the examples of primary calibration algorithms for hemoglobin discussed below, the pixel number could be used instead of the actual wavelength. The actual wavelength is required when the wavelength corresponding to a particular pixel in one linear diode array is different from the wavelength corresponding to the same pixel in a different linear diode array. Also, the actual wavelength is useful for identifying analytes with characteristic absorbance peaks. For example, hemoglobin has two characteristic absorbance peaks at 540nm and 578nm. These absorbance peaks may be used to identify the hemoglobin signal. If one wavelength is used to perform the wavelength calibration, it is preferred that a wavelength that is close to the primary calibration wavelength of the primary calibration algorithm be used. If more than one wavelength is used, it is preferred that the wavelengths are significant with respect to the analyte signal. For example, a calibration algorithm for hemoglobin would preferably comprise a first beam of about 593nm to 608nm, according to the hemoglobin primary calibration algorithm in Example 2.

25

In the examples described below, the apparatus are used to measure concentrations of interferents in serum and plasma using radiation, most preferably using near infrared radiation and the adjacent visible radiation spectrum, for example, but not limited to, any suitable wavelength from about

300nm to about 2500nm, preferably from about 500nm to about 1100nm, and more preferably, from about 500nm to about 800nm.

B) PRIMARY CALIBRATION

5 The process of primary calibration is used to develop a primary calibration algorithm for each analyte. Primary calibration of an apparatus is a cumbersome, time intensive exercise because the primary calibration set is relatively large, and the samples in the primary calibration set should be real or very close to real samples. Preferably, samples include all the absorbance
10 variability expected in a sample, whereby the sample variability becomes built into the primary calibration algorithm. Vessels also contribute variability, and it is possible to develop one or more primary calibration algorithm using a combination of more than one vessel, whereby the vessel variability may become built into the primary calibration algorithm. However, development of
15 primary calibration algorithms that are specific to a particular type of vessel is preferred.

 A primary calibration algorithm can be obtained as follows:
Absorbance spectra are obtained for several samples each of which having a
20 different concentration of a given analyte for which the primary calibration algorithm is being developed. It is preferred that the samples include all the absorbance variability expected in a sample, whereby the sample variability becomes built into the primary calibration algorithm. A multiple linear regression is then performed to determine at which wavelengths the
25 absorbance (or derivative of absorbance) correlates best with the concentration of the analyte. Then, using multiple linear regression a linear combination having the absorbance, or derivative of absorbance, at specific wavelengths as the independent variable and the concentration of the analyte as the dependent variable is generated.

Software tools used for developing primary calibration algorithms comprises of the following: Mathlab™ used to create programs for smoothing absorbances and derivative of absorbances. MS Excel™ may be used to develop macros for calculating derivative of absorbances; StatView™ used to
5 create algorithms by a process called "step-wise multiple linear regression." In the step-wise linear regression, absorbance or derivative of absorbance measurements for all the wavelengths may be presented to the StatView™ program. Only the wavelengths at which the absorbance or derivative of absorbance exhibits a correlation coefficient within a predetermined range
10 are selected for the algorithms. Pirouette™ may be used to create calibration algorithms by Partial Least Squares (PLS) or Principal Component Analysis (PCA), using the measurements for all the wavelengths, or selected sections of the absorbance spectra. It will be appreciated however that other software tools may also be used.

15

Each apparatus used for the purpose of determining the concentration of analytes in a sample, must be calibrated according to this process on an as needed basis. The primary calibration procedure, in respect of interferences, is set out here. Although it will be appreciated by those skilled in
20 the art that the procedures set out here for interferences will apply to any analyte, in any sample, and a primary calibration algorithm may contain from a single wavelength term, in the simplest case, to multiple terms that use all the wavelengths.

25

Installation of a primary calibration algorithm in a second apparatus can take place at any time, including when the second apparatus is being manufactured, or just prior to use. Preferably, the primary calibration algorithm(s) is/are incorporated into the hardware of a second apparatus, for example, but not limited to, an EPROM that is installed in a second apparatus
30 thereby incorporating the primary calibration algorithm. However, the primary

calibration algorithm may also be stored on any other suitable medium, for example, but not limited to a calibration diskette and incorporated in the apparatus as required. The incorporated algorithms are preferably used in the second apparatus after the absorbance measurements are interpolated and mapped onto the same standard set of wavelengths used for the primary calibration and after photometric correction is performed for each relevant wavelength. Preferably, calibrators used for photometric correction are synthetic fluids, but other materials may also be used. More preferably, a calibration package as discussed under the title "Calibration Package" comprising calibrators and at least one primary calibration algorithm may be used. Examples of slopes and intercepts which may be used for photometric corrections are given in Table 1 of Example 1. The wavelengths in Example 1 are from a standard set of wavelengths.

Examples of deriving a primary calibration algorithm for hemoglobin, bilirubin and turbidity are provided below, however, it should be understood that these examples are not to be considered limiting in any manner.

Calibration sets for hemoglobin (Hb), biliverdin (BV), bilirubin (BR) and turbidity in serum are described below. Intralipid™ (IL), a fat emulsion, is used to mimic turbidity. The four analytes mentioned are the major contributors of absorbance variability in serum and plasma samples. Following the description of the calibration sets, are several sample primary calibration algorithms.

25

Hemoglobin and Biliverdin

To prepare a Primary Calibration Algorithm for hemoglobin, sixty serum specimens with no visible interferents were stored refrigerated or frozen until used. More or fewer specimens may be used so long as a sufficient number is used to provide robust algorithm(s). Hb, IL, BR and BV were added to the

30

normal sera to give final concentrations of 0-6.1 g/L, 0-5.1 g/l, 0-42.7 mg/dL, and 0-4.4 mg/dL respectively. Stock Hb was prepared by replacing the plasma (must be free from all interferents) from a blood sample, with twice its volume of water, and lysing the cells through three freeze-thaw cycles. For each cycle the blood was left in the freezer for 45 – 60 minutes, and then removed and placed on a rocker at room temperature for 30 - 45 minutes. Hb content of the lysate was measured by the reference method described below, after removing the red blood cell (RBC) debris and unlysed RBC's by centrifuging at 10,000 x g for 10 minutes. Any method which provides a reliable determination of Hb content may be used. A typical hemolysate contains approximately 100 g/L Hb. CO-oximetry suggests that more than 95% of the Hb is in the oxy-Hb state. Stock BV was prepared by dissolving biliverdin dihydrochloride (from Sigma) initially in 50% methanol-50% water, and diluting further with phosphate buffered saline (PBS). Stock IL also known as Travamulsion™ (from Clintec-Nestle & Baxter) has a concentration of 10%. Stock BR was prepared by dissolving Ditauro-Bilirubin (from Porphyrin Products, Logan, Utah, USA) in interferent-free serum, to a concentration of 500 mg/dL. The spectral absorbance data were recorded for the 60 samples using different polypropylene dispensing tips. Out of the 60 samples, odd numbers were used for the calibration set, and even numbers were used as the validation set.

Bilirubin

The sample set used for Hb and BV calibrations are not typically used for BR calibration, because the absorbance due to either Hb >4g/L or IL >4g/L, approaches the limit of the apparatus in the region around 524nm, a primary wavelength used for BR calibration. Instead, a separate set of 60 samples were prepared and tested. As will be readily appreciated by those skilled in the art, the sample set used for primary calibration should be of a size sufficient to include most of the variability encountered with actual patient

samples, such as serum or plasma. The samples were prepared as before by adding Hb, IL BR and BV to the normal sera to give final concentrations of 0-2.6 g/L, 0-3.6 g/l, 0-37 mg/dL, and 0-4.4 mg/dL respectively. The spectral absorbance data were recorded for the 60 samples using different polypropylene dispensing tips. Out of the 60 samples, odd numbers were used for the calibration set, and even numbers were used as the validation set. The stock interferents were prepared as described above for Hb, and the BR concentrations were adjusted by the factor 1.23. The 1.23 factor that was derived previously from the slope of the regression line obtained from a validation set using real icteric serum and plasma samples.

Turbidity

Turbidity in serum and plasma is caused mainly by the presence of fat particles, particularly chylomicrons. Intralipid (IL) is a fat emulsion which mimics naturally-occurring chylomicrons, and therefore may preferably be used to simulate turbidity in serum and plasma.

Samples used for Hb and BR calibration are preferably not used for IL calibration because the Hb stock solution contributes significant light scattering due to unlysed red blood cells (RBC's) and RBC fragments. Centrifugation of the hemolysate was unable to remove all the unlysed RBC and RBC fragments.

Forty samples of PBS (phosphate buffered saline) were spiked with 10% Intralipid to produce concentrations of 0-20 g/L. The spectral absorbance data were recorded for the 40 samples using different polypropylene dispensing tips. Out of the 40 samples, the odd numbers were used for the calibration set, and the even numbers were used as the validation set.

As mentioned above, the primary calibration described herein is exemplary of the work involved in developing primary calibration algorithms. Other analytes may be used to develop such primary calibration algorithms.

5 Primary Calibration Algorithms

The following Primary Calibration Algorithms were developed for Hb, BV, BR and IL (turbidity) in disposable polypropylene dispensing tips using the primary calibration approach set forth above.

10 Hemoglobin

$\text{g/L Hb} = 16.81(1\text{st D A}584) - 79.47(1\text{st D A}599) + 60.95(1\text{st D A}617) + 0.24$
where (1st D A) is the first derivative of the absorbance measurement at the wavelength specified in nanometers.

15 Bilirubin

$\text{mg/dL BR} = -293.1(1\text{st D A}524) - 327.8(1\text{st D A}587) + 451.7(1\text{st D A}602) - 7.5$

where (1st D A) is the first derivative of the absorbance measurement at the wavelength specified in nanometers:

20

Turbidity

Turbidity is measured in terms of equivalent IL concentration.

$\text{In (g/L IL)} = 1.867(\text{A}700) - 0.447(\text{A}700)^2 + 0.041(\text{A}700)^3 - 1.33$

where (A) is the raw absorbance measurement at the wavelength specified in
25 nanometers.

A further set of primary calibration algorithms developed on another apparatus used to test the samples for measurement in disposable polypropylene dispensing tips, are as follows:

30

Hemoglobin

$$(\text{g/L}) \text{ Hb} = 30.14 (1\text{st D A}591\text{nm}) - 27.98 (1\text{st D A}610\text{nm})$$

where (1st D A) is the first derivative of the absorbance measurement at the wavelength specified.

5

Bilirubin

$$\text{mg/dL BR} = 142.09 (1\text{st D A}511\text{nm}) + 89.9 (1\text{st D A}554\text{nm}) - 4.47$$

where (1st D A) is the first derivative of the absorbance measurement at the wavelength specified.

10

Biliverdin

$$\text{mg/dL BV} = 160.29 (1\text{st D A}718\text{nm}) - 206.15 (1\text{st D A}781\text{nm}) + 1.42$$

where (1st D A) is the first derivative of the absorbance measurement at the wavelength specified.

15

Turbidity

$$\text{g/L IL} = 296.01 (\text{A}900\text{nm}) - 0.04$$

where (A) is the raw absorbance measurement at the wavelength specified.

20

The Primary Calibration Algorithms referred to herein are non-limiting examples obtained by a process of step-wise multiple linear regression. Other methods like partial least squares (PLS) and principle components analysis (PCA) may also be used.

25

It is noted that in developing a primary calibration algorithm the absorbance and wavelength variabilities acceptable in a batch of apparatus depends on the accuracy requirement for the analyte. Further, the absorbance and wavelength variabilities acceptable in a batch of apparatus depend on the requirement of the inter-apparatus variability of the analyte

30

concentration.

As indicated above, several examples for primary calibration algorithms for the same analyte have been shown. Some reasons why different primary calibration algorithms are generated are as follows: the primary calibration algorithms are generated by different people, using the same software tools; the primary calibration algorithms are generated by the same or different people, using different software tools; wider analytical range is required (e.g., measuring high levels of Hb-based blood substitutes, which give the appearance of high levels of real Hb), as described in WO 9839634 (which is incorporated herein by reference).

The conventional method to deal with a concentration that exceeds the upper limit of the analytical range is to dilute the sample. New wavelengths may be used, and/or new mathematical transformation(s) may be used, instead of diluting the sample. For example, either a cubic primary calibration algorithm that uses natural logs, or a linear single term primary calibration algorithm was shown for turbidity; said cubic primary calibration algorithm was developed for a wider analytical range of turbidity. As another example, greater accuracy may be required at low concentrations of Hb and therefore a calibration algorithm may be developed to measure traces of Hb in serum or plasma over a small analytical range, and when the predicted value exceeds the upper limit of said analytical range, the software could be prompted to use a different calibration algorithm, which covers a wider analytical range. As yet another example, when the wavelength range provided by the apparatus is limited, for example, one apparatus may only have the capability of measuring 450 - 750nm, instead of 450 - 1100nm, the primary calibration algorithms for IL may use 700nm and 900nm respectively.

An important aspect of primary calibration algorithm development is the validation process that is described next. Validation requires a sample set

similar to the calibration set; the calibration set cannot be used for validation. Validation may be a test of robustness of the primary calibration algorithms, depending on the overall variability presented in the validation set. If the calibration algorithm for a particular analyte cannot accurately predict the concentration of the analyte in the sample, it could be that the variability in the test sample was not included in the primary calibration set. In other words, the primary calibration algorithm for said analyte lacked robustness.

C) VALIDATION

10 Hemoglobin and Biliverdin

As mentioned above in the section on primary calibration, out of the 60 samples, odd numbers were used for the primary calibration set, and even numbers were used as the validation set.

15 Bilirubin

As mentioned above in the section on primary calibration, out of the 60 samples, odd numbers were used for the primary calibration set, and even numbers were used as the validation set.

20 Turbidity

As mentioned above in the section on primary calibration, out of the 40 samples, the odd numbers were used for the primary calibration set, and the even numbers were used as the validation set.

25 An important requirement of validation is the availability of a reliable reference method for a particular analyte. A reference method is required to determine the actual or true concentration of the analyte in the sample. The difference in the true concentration, as measured by the reference method, and the concentration predicted by the primary calibration algorithm, is a

measure of prediction accuracy of the primary calibration algorithm. The next section describes some reference methods.

Reference Methods

5 The sample absorbance data and the concentrations of the interferents are required for development of calibration algorithms for the interferents. Any errors in the reference methods used to measure the concentration of the interferents will affect the prediction accuracy of the primary calibration algorithms.

10

Hemoglobin

For validation using spiked serum samples, the Hb concentration in the hemolysate was used to calculate the reference Hb values. For validation using serum or plasma samples, methods known to those skilled in the art, 15 such as for example: Tietz Textbook of Clinical Chemistry, 1994, page 2024, may be used; the same method is used to determine the Hb concentration in the hemolysate.

For accurate hemoglobin measurement in serum and plasma samples, 20 preferably they will be the only analyte present, and the absorbances at 578nm, 562nm, and 598nm may be measured and the concentration of hemoglobin determined using the following equation:

$$\text{mg/dL Oxy-hemoglobin} = 155.0 A_{578} - 86.1 A_{562} - 68.9 A_{598}$$

25

where: A is the absorbance at the wavelengths specified in nanometers. Appropriate dilutions are performed to produce absorbances between 1.0 and 2.0, for the 578nm wavelength.

As appropriate, dilutions may be made with 10mg/dL sodium carbonate. Any other reliable reference method may be used. For the present illustration of primary calibration, a Shimadzu single-beam scanning spectrometer with a grating was used, although any comparable apparatus may be used.

5

Biliverdin

The weighed amount of added biliverdin dihydrochloride was used instead of an actual method for measuring the concentration of BV in the samples.

10

Bilirubin

For validation using spiked serum samples, the concentration of synthetic ditauro- bilirubin added, divided by 1.23, was used as the reference total BR value. For validation using patient serum or plasma samples, any total
15 bilirubin test performed by chemistry analyzers (e.g., Beckman-Coulter or Ortho-Clinical Diagnostics general chemistry analyzers) used by accredited medical laboratories could be used as the reference method.

As mentioned before, the 1.23 factor used was derived previously from
20 the slope of the regression line obtained from a validation set using real icteric serum and plasma samples.

Turbidity

For validation using PBS spiked with IL, the concentration of the IL
25 present was used as the reference value. In Tietz Textbook of Clinical Chemistry, 1994, page 2022, absorbance measurement at 700nm is used to make corrections for turbidity. Therefore, for this instance, validation using patient serum or plasma samples, absolute absorbance at 700nm may be adopted as the reference method. A commercial single beam
30 spectrophotometer should be used with a 10mm x 10mm cuvette, referenced

against PBS. The absorbances before or after dilutions should preferably be between 0.2 and 2.0. Photometric and wavelength accuracy of the reference spectrophotometer should preferably be demonstrated before use.

5 D) CALIBRATION ALGORITHM TRANSFER

The present invention provides a method for transferring a calibration algorithm from a first apparatus to a second apparatus whereby the second apparatus need not be calibrated in the same way in which the first apparatus was calibrated, that is by conducting a primary calibration.

10

According to one embodiment the present invention provides a method for calibration algorithm transfer from a first apparatus to a second apparatus comprising:

(i) obtaining a first set of absorbance measurements of a set of
15 calibrators on a First Apparatus that is in control at wavelengths from a first wavelength calibration table;

(ii) establishing a second wavelength calibration table on a second apparatus, the first and the second wavelength calibration table may be the same or different, and obtaining a second set of absorbance measurements
20 of the set of calibrators on the Second Apparatus, at wavelengths from the second wavelength calibration table;

(iii) determining a first interpolated absorbance for the first absorbance measurements for at least one wavelength of a Standard Set of Wavelengths, and determining a second interpolated absorbance for the
25 second absorbance measurements for the at least one wavelength of the Standard Set of Wavelengths,

(iv) deriving a First Linear Regression Equation for each of the at least one wavelength of the Standard Set of Wavelengths using the first and the second interpolated absorbance measurement;

(v) incorporating the First Linear Regression Equation and at least one Primary Calibration Algorithm onto the Second Apparatus.

Subsequent to calibration algorithm transfer, according to an embodiment of the invention, there is further provided a method of determining the concentration of an analyte in a sample in a second apparatus comprising:

- (a) performing a Calibration Algorithm Transfer according to the method described above;
- (b) measuring an absorbance of the sample on the second apparatus, and determining a sample interpolated absorbance for at least one wavelength of the Standard Set of wavelengths;
- (c) adjusting the interpolated absorbance with the First Linear Regression Equation to obtain an Adjusted Interpolated Absorbance; and
- (d) calculating a concentration for the analyte by applying the at least one Primary Calibration Algorithm for the analyte to the Adjusted Interpolated Absorbance.

The first linear regression equation is derived from a plot of interpolated absorbance measurements, the first interpolated absorbance measurements preferably on an x-axis, the second interpolated absorbance measurements preferably on a y-axis, and the first linear regression equation having a y-intercept and a slope. It should be understood that the first interpolated absorbances may be plotted on the y-axis and the second interpolated absorbances plotted on the x-axis. In such case, Adjusted Interpolated Absorbance = (Slope x Interpolated Absorbance) + Y-intercept. By interpolated absorbance, it is meant the absorbance value for a specific wavelength of the standard set of wavelengths. If the specific wavelength of the standard set of wavelengths is the same as the wavelength already assigned to a pixel of a linear diode array, then the interpolated absorbance is the same as the measured absorbance. However, if the specific

wavelength is different from the wavelength assigned to a pixel, then the value of interpolated absorbance, at the specific wavelength, is estimated from the measured absorbance values of at least two adjacent pixels whose wavelengths encompass one or more standard wavelengths.

5

Examples of first linear regression equations are shown in Example 1. This adjustment of absorbance is also referred to as photometric correction. Accordingly, the measurement made in respect of the sample once adjusted provides a result comparable to that which would be obtained if the sample
10 had been measured on the first apparatus. In this way, the second apparatus need not be calibrated through primary calibration as was done for the first apparatus.

According to a preferred embodiment the primary calibration
15 algorithms and the interpolated absorbance measurements of the calibrators made on the first apparatus are electronically stored and installed on the second apparatus. Electronic storage includes more preferably a floppy diskette or CD. More preferably, a calibration package as discussed under the title "Calibration Package" comprising of calibrators and at least one
20 primary calibration algorithm may be used.

It is also preferred that the samples used for the Primary Calibration, and the calibrators, are placed within like vessels having optical properties substantially similar. Non-limiting examples of vessels that can be used are, a
25 pipette tip, a labeled test tube, an unlabeled test tube, a blood bag tubing, any transparent container, or any translucent container. In one aspect of the invention the sample can be liquid for example, a biological fluid, including but not limited to bodily fluids for example serum or plasma, or non-biological, for example but not limited to a beverage or a synthetic calibrator, or solid for
30 example cheese, which is not to be considered limiting in any manner, or a

medicament, for example a pharmaceutical tablet. In another aspect of the invention, a sample exhibits one or more properties that may be determined spectrophotometrically. A sample typically comprises one or more analytes. By analyte it is meant a substance being measured in a sample. A sample is
5 characterized in that the energy of one or more wavelengths of light can be either transmitted, absorbed, scattered, reflected, or a combination thereof, by an analyte within the sample. The light that is measured in the preferred embodiment is the transmitted light, and said transmitted light is converted into absorbance units. It should be understood that said absorbance units
10 are not true absorbance units, as would be obvious by those skilled in the art.

In another aspect of the invention, a set of calibrators are defined as two or more calibrators and said set of calibrators and said sample are placed within a like vessel used for the primary calibration. Furthermore, any type of
15 calibrator suitable for producing a set of first linear regression equations based on interpolated absorbances obtained using a first apparatus and a second apparatus, may be used. For example, a set of calibrators may comprise translucent substances such as fiberglass which may vary in thickness, color, or finish, depending on the type of analyte to be measured.
20 The calibrators may also comprise solid particles in liquid suspension, e.g., polystyrene beads in a buffer of similar density as the polystyrene beads. Calibrators may also comprise fat emulsion, for example fat particles in suspension. Calibrators may also comprise colored glass of varying thickness and finish. In any event, the calibrators should produce a suitable
25 set of first linear regression equations on a First Apparatus and a Second Apparatus based on interpolated absorbances on the respective apparatus. In the preferred embodiment, the set of calibrators tested on both first apparatus and second apparatus, are from the same batch. Further, the calibrators used on said second apparatus may be the same as said set of
30 calibrators used on said first apparatus.

In another aspect of the invention, a standard set of wavelengths comprise wavelengths from about 300nm to about 2500 nm, preferably from about 500nm to about 1100nm, and more preferably, from about 500nm to
5 about 800nm.

In another aspect of the invention, the first and said second absorbance measurements are determined one or more times at each wavelength of a calibration table, and an average of said first, or said second,
10 interpolated absorbance(s) is mapped to said standard set of wavelengths. Further, said standard set of wavelengths used by both said first apparatus and said second apparatus are the same.

According to one embodiment of the invention, a standard set of
15 wavelengths is a set of wavelengths used by all apparatus in conjunction with the apparatus-specific wavelength calibration table, used to generate interpolated absorbances from the actual absorbances. More details are provided in the section entitled "Wavelength Calibration." The actual absorbances of a sample tested on an apparatus are measured at
20 wavelengths from the wavelength calibration table, and these actual absorbances are interpolated and mapped onto the standard set of wavelengths. The primary calibration algorithm(s) is/are preferably applied to the interpolated mapped absorbances - not the actual absorbances. Without wishing to be limiting in any manner, an example of a standard set of
25 wavelengths includes 500 to 1100nm, in increments of 2nm, i.e., a pixel dispersion of 2nm. However, other wavelength ranges and pixel dispersions may be used as required, and as would be known by one of skill in the art. Spectrophotometers may be designed to have the same wavelength calibration table by projecting a single beam of known
30 wavelength on the same pixel in each linear diode array, and extrapolating to

obtain the wavelengths corresponding to the other pixels using the same pixeldispersion. When all apparatus use the same wavelength calibration table, such that the wavelength calibration table becomes the standard set of wavelengths, the step of interpolating absorbances is automatically
5 eliminated. As previously mentioned, pixeldispersion means the wavelengths encompassed by two adjacent pixels of a linear diode array, usually measured in nanometers (nm) per pixel. For example, if two lasers of 600nm and 900nm are used for wavelength calibration, and they are projected on pixel 20 and pixel 220 respectively, that means 300nm (i.e., $900 - 600\text{nm}$) are
10 encompassed by 200 pixels (i.e., $220 - 20$ pixels). Therefore the pixeldispersion is calculated to be 1.5nm per pixel (i.e., 300nm divided by 200 pixels).

According to one aspect of the invention, the pixeldispersion may be
15 predetermined by the steps comprising:

- (i) Projecting an electromagnetic radiation of known wavelength, onto a first pixel preferably towards an end of a linear diode array;
- (ii) Using a second electromagnetic radiation of a known wavelength, projecting the second laser or second light onto a second pixel
20 preferably at the opposite end of the linear diode array;
- (iii) Identifying the first and second pixels within the linear diode array;
- (iv) Calculating the pixeldispersion in nanometers per pixel or any wavelength units per pixel as described above; and

25

In one aspect of the invention, the pixeldispersion may be calculated in step (iv) for one linear diode array and may be used as the predetermined pixeldispersion. In another aspect of the invention, steps (i) to (iv) may be repeated on one or more apparatus, preferably, the apparatus have similar
30 linear diode arrays, and the average pixeldispersion may be used as the

predetermined pixeldispersion. It is understood that the predetermined pixeldispersion may be provided by the manufacturer of the spectrophotometer.

5 The wavelength calibration table may be obtained by:

- (i) projecting a first electromagnetic radiation of known wavelength, onto a first pixel of a first linear diode array of said first apparatus, or a second linear diode array of said second apparatus;
- (ii) using a second electromagnetic radiation of known wavelength,
10 said second electromagnetic radiation having a different wavelength than said first electromagnetic radiation, projecting said second electromagnetic radiation onto a second pixel of said first or said second linear diode array;
- (iii) identifying said first and second pixels within said first or said second linear diode array;
- 15 (iv) calculating a pixeldispersion for said first or said second linear diode array; and
- (v) assigning a wavelength to each pixel within said first or said second linear diode array to produce said wavelength calibration table using said pixeldispersion and either said first electromagnetic radiation of known
20 wavelength, and said first pixel, or said second electromagnetic radiation of known wavelength and said second pixel.

In another aspect of the invention, the wavelength calibration table is obtained by projecting only the first electromagnetic radiation of known
25 wavelength onto an identified first pixel of said linear diode array, using a predetermined pixeldispersion. Further, according to this method using a only a first beam, wherein said method is repeated on a second apparatus, preferably with similar components, said wavelength calibration table of said second apparatus is determined using the same pixel number of said first

pixel. Therefore, the wavelength calibration table could be the standard set of wavelengths.

In a further aspect, the wavelength calibration table of the second
5 apparatus may be determined by projecting only a first beam on a pixel of a different pixel number as that of the first apparatus.

In a preferred embodiment where analyte accuracy is critical, two
wavelengths should be used to facilitate determination of the pixel dispersion
10 for the apparatus, and a specific wavelength calibration table is generated for each apparatus using the two wavelengths, their corresponding pixel numbers, and the calculated pixel dispersion. In another embodiment where analyte accuracy is not as critical, a single wavelength, its corresponding pixel number, and a predetermined pixel dispersion should be used to generate a
15 wavelength calibration table. It is understood that if the single wavelength is projected onto the same pixel number of like linear diode arrays in different apparatus, the same wavelength calibration table would be generated. It is also understood that if the same wavelength calibration table is generated, said wavelength calibration table becomes the standard set of wavelengths,
20 thereby eliminating the step of wavelength interpolation and mapping. It is also understood that if the wavelength calibration table is the same for each apparatus, pixel numbers can be used in the primary calibration algorithm(s) instead of wavelengths. Example 6 illustrates the resulting wavelength inaccuracy when the beam does not project on the same pixel number in
25 every linear diode array. It is obvious to those skilled in the art that wavelength inaccuracy is translated into absorbance inaccuracy.

E) RECALIBRATION

The present invention provides a method for recalibrating an apparatus that was previously calibrated by calibration algorithm transfer, but is no longer in control, comprising the following steps:

- 5 (i) obtaining absorbance measurements of a set of calibrators on said apparatus, said set of calibrators having assigned absorbance values, said apparatus comprising a Primary Calibration Algorithm;
- (ii) determining interpolated absorbance values for said absorbance measurements for at least one wavelength of a Standard Set of
10 Wavelengths;
- (iii) establishing a Second Linear Regression Equation in said apparatus, using said interpolated absorbance values and said assigned absorbance values; and
- (iv) incorporating said Second Linear Regression Equation on said
15 apparatus to produce a recalibrated apparatus.

Subsequent to recalibration, according to an embodiment of the invention, there is further provided a method of determining the concentration of an analyte in a sample in a Second Apparatus comprising the following
20 steps:

- (a) recalibrating said apparatus according to the method described above;
- (b) measuring an absorbance measurement of said sample;
- 25 (c) deriving an interpolated absorbance for said absorbance measurement for at least one wavelength of said Standard Set of Wavelengths in said recalibrated apparatus;
- (d) adjusting said interpolated absorbance measurement with said Second Linear Regression Equation to obtain an Adjusted Interpolated
30 Absorbance; and

(e) calculating a concentration for said analyte by applying said Primary Calibration Algorithm for said analyte to said Adjusted Interpolated Absorbance.

5 It is understood that the definition of terms used in recalibration are as described at the beginning of "Detailed Description of Invention." Further, it should be understood that recalibration is a sequel to calibration algorithm transfer, when an apparatus is no longer in control.

10 Second linear regression equations are developed for each relevant wavelength of a standard set of wavelengths, with the interpolated absorbance measurements from the Set of calibrators versus absorbance measurements assigned to the lot or batch of calibrators, after the absorbances of said calibrators were measured on the first apparatus when it
15 was in control. The second linear regression equation is derived from a plot of the absorbance measurements of the calibrators, preferably with the interpolated absorbance from the second apparatus on a y-axis and the assigned absorbance from the first apparatus on an x-axis. Each generated second linear regression equation having an intercept and slope, is then
20 stored electronically in the apparatus being recalibrated. Electronic storage includes more preferably a floppy diskette or CD. Accordingly, the interpolated absorbance in respect of the sample once adjusted, provides a result comparable to that which would be obtained if the sample had been
25 measured on the first apparatus. In this way, the second apparatus need not be calibrated through a primary calibration as was done for the first apparatus.

 For clarity, the term "Second Linear Regression" was used for recalibration (and also for calibration as will be seen under the title
30 "Calibration") of an apparatus and the term "First Linear Regression" was

used for calibration algorithm transfer. According to an aspect of the invention, when compared to calibration algorithm transfer the only new requirement in the process of recalibration of an apparatus is determination of interpolated absorbances of two or more calibrators for at least one wavelength of a standard set of wavelengths on said apparatus. Further, the first apparatus is not required, because recalibration is a sequel to calibration algorithm transfer, and therefore the primary calibration algorithm(s) is, or are, already incorporated in the second apparatus.

10 F) CALIBRATION

Calibration is the process of establishing and incorporating one or more second linear regression equations in an apparatus that was never subjected to calibration algorithm transfer. Therefore the primary calibration algorithm(s) is or are yet to be incorporated in said apparatus. Like recalibration, calibration does not require the first apparatus. A method for calibrating an apparatus that was never subjected to calibration algorithm transfer, comprises the following:

- (i) obtaining absorbance measurements of a Set of Calibrators on said apparatus, said apparatus lacking a primary calibration algorithm, and said set of calibrators having assigned absorbance values,
- (ii) determining interpolated absorbance values for said absorbance measurements for at least one wavelength of a Standard Set of Wavelengths;
- (iii) establishing a Second Linear Regression Equation in said apparatus, using said interpolated absorbance measurements and said assigned absorbance values; and
- (iv) incorporating said Second Linear Regression Equation, and at least one Primary Calibration Algorithm on said apparatus, to produce a calibrated apparatus.

Subsequent to calibration, according to an embodiment of the invention, there is further provided a method of determining the concentration of an analyte in a sample in a calibrated apparatus comprising:

- 5 (a) calibrating said apparatus according to the method described above;
- (b) measuring an absorbance value of said sample;
- (c) deriving an interpolated absorbance from said absorbance value for at least one wavelength of said Standard Set of Wavelengths in said calibrated apparatus;
- 10 (d) adjusting said interpolated absorbance measurement with said Second Linear Regression Equation to obtain an Adjusted Interpolated Absorbance; and
- (e) calculating a concentration for said analyte by applying said Primary Calibration Algorithm for said analyte to said Adjusted Interpolated
- 15 Absorbance.

It should be understood from the definitions of terms used in "Calibration" are as described at the beginning of "Detailed Description of Invention." Further, it should be understood that calibration is for an apparatus that does not have any relevant primary calibration algorithm incorporated in said apparatus.

The essential components of a calibration package (discussed under the title "Calibration Package") comprises of:

- 25 i) At least one primary calibration algorithm;
- ii) At least two calibrators; and
- iii) Assigned absorbances for said calibrators.

The assigned absorbances for calibrators are preferably obtained from the first apparatus when said apparatus is in control. By definition of first apparatus, the primary calibration algorithm(s) is/are developed for said first

30

apparatus. Therefore, the assigned absorbances of calibrators and the primary calibration algorithm(s) should preferably originate from the same apparatus. During recalibration, if the calibrators' assigned absorbances do not originate from the same apparatus from which the primary calibration
5 algorithm(s) originated, there exists a risk of having a mismatch between the calibrators' assigned absorbances and the primary calibration algorithm(s). In order to mitigate this risk, it is preferred that the appropriate primary calibration algorithm(s) be reincorporated in the apparatus. It becomes obvious that this preferred form of recalibration is no different from the
10 process of calibration. Therefore, in the preferred embodiment, calibration is a safer approach than recalibration. In other words, calibration is preferred to recalibration when a calibration package is available.

15 G) CALIBRATION OF A SECOND APPARATUS USING DERIVATIVE OF ABSORBANCE

The inventor has also found that the process of determination of analyte concentration can be accomplished by using an order of derivative of the absorbance in the primary calibration algorithm, where photometric correction may or may not be necessary, provided that the order of derivative
20 of absorbance used in the primary calibration algorithm at the selected wavelength(s) does not contain significant inter-apparatus variability, although the interpolated absorbances of the same sample for at least one of a standard set of wavelengths may contain significant variability. It is well known that absorbance variability can be removed in certain sections of the
25 absorbance spectra, by using an order of derivative of the absorbance. The acceptable variability in the derivative of the absorbance(s) at the wavelength(s) chosen for the calibration depends upon the inter-apparatus variability in the predicted concentration of the analyte.

According to this approach while the first derivative of absorbance removes most variability between a first and second apparatus, higher derivatives, while still useful and within the scope of the present invention, tend to introduce sharper peaks and accordingly are not as preferred as the first derivatives. Derivatives of absorbance beyond the 2nd increase the complexity of the spectra and magnify the noise i.e., the signal to noise ratio decreases. Because the peaks are sharper with the higher order of derivatives, wavelength accuracy becomes more critical for accurate determination of analyte concentration.

10

As the requirement of analyte prediction accuracy increases, and as the complexity of the calibration algorithm increases, the need for photometric correction increases. Also, wavelength accuracy requirement depends on the required accuracy of the analyte. For example, attempts can be made to make the instruments with the same wavelength calibration table, thereby eliminating the need for interpolating and mapping absorbances to a standard set of wavelengths. Methods that may be used to generate a standard set of wavelengths are discussed in details elsewhere, for example, in the section on "Wavelength Calibration."

20

Wavelength accuracy is still critical for efficient calibration algorithm transfer, when the derivatives of absorbance are used, without performing photometric adjustments. To illustrate this point, wavelength inaccuracies were added to the first and second derivative of absorbance obtained from a second apparatus by shifting the first and second derivative of absorbance data to the right by two pixels (2 columns). The effect of doing this may be seen in Figure 17 and Figure 18 for the second and first derivative respectively. The offsets are obvious due to the wavelength inaccuracy in the second apparatus. These offsets can be minimized by interpolating and mapping the absorbances to a standard set of wavelengths as described

30

above, without performing photometric adjustments. Further improvement will be seen if the wavelength calibration of each spectrometer is performed with two wavelengths, preferably provided by lasers as opposed to filters, and preferably narrow band-pass filters if filters are used.

5

Methods disclosed herein and used for spectrophotometer wavelength calibration, the use of a standard set of wavelength, and the use of photometric adjustment using synthetic calibrators, all contribute to efficient calibration algorithm transfer. Proper wavelength calibration requires skill and time. One may choose the best method for spectrophotometer wavelength calibration and omit the photometric adjustment. Alternatively, one may compromise on the wavelength calibration and convert the absorbances measured to absorbances at a standard set of wavelengths, followed by photometric adjustment. Photometric adjustment could be easily implemented by software, following testing of at least two synthetic calibrators. The methods applied to transfer calibration algorithms depend on the required prediction accuracy of analytes. Note that the accuracy referred to here means closeness to the predictions by a first apparatus.

20 As mentioned, an aspect of calibration without photometric correction is the selection of specific wavelengths, i.e., wavelength optimization.

For multiple linear regression, a method used to select wavelengths that show minimum variability in the absorbance, or first derivative of absorbance (or any derivative of absorbance) is to test the primary calibration samples on several apparatus that have the expected amount of absorbance and wavelength variability, and use the data from all the apparatus in the primary calibration set. After the wavelength(s) is/are selected, only the data generated from a first apparatus should be used to develop the primary calibration equation, using the predetermined wavelength(s).

30

By "Derivative of Absorbance" it is implied an order of derivative of the absorbance spectrum. For example, the first order derivative of absorbance at a particular wavelength is the slope of the absorbance spectrum at said wavelength; the second order derivative of absorbance at a particular wavelength is the slope of the first derivative absorbance spectrum at said wavelength, and so on. Methods of calculating a derivative of absorbance at a particular wavelength is well known by those skilled in the art. The first derivative of absorbance at a particular wavelength is the simplest calculation and it may be as simple as the difference in absorbances at the two wavelengths that encompass said particular wavelength. Other methods of calculating derivative of absorbance may use the absorbances at several different wavelengths, where smoothing of the results is also achieved, i.e., removal of spikes also called noise. The more wavelengths used, the greater the degree of smoothing, and also the greater the degree of smoothing, the greater the loss of signal details in the spectrum. The minimum number of wavelengths that may be used to calculate a derivative of absorbance is two wavelengths; the order of that derivative is first order.

According to one aspect of the invention, a method of determining the concentration of an analyte in a sample in a second apparatus comprises the following:

- (i) incorporating at least one primary calibration algorithm that uses an order of derivative of absorbance obtained for at least one wavelength of a standard set of wavelengths, on said second apparatus;
- (ii) measuring absorbance values of said sample at three or more wavelengths from a wavelength calibration table on said second apparatus;
- (iii) determining interpolated absorbance values from said absorbance values for wavelengths from a standard set of wavelengths;

(iv) obtaining a derivative of said interpolated absorbance values, using said order of derivative; and

(v) calculating a concentration of said Analyte in said sample, by applying said Primary Calibration Algorithm to said derivative.

5

In the case where the wavelengths of said standard set of wavelengths are the same as wavelengths of the wavelength calibration table, two or more wavelengths are used in the step of measuring (step (ii)) since this is sufficient to determine the derivative.

10

In the above method the derivative can be of a first, second, or third order. Also the primary calibration algorithm may comprise 1, 2, 3, or 4 wavelengths.

15

In another aspect of the invention, the at least one wavelength is optimized and derived from at least two or more apparatus. Optimization comprises visual examination of the derivative of absorbance spectra, and choosing wavelength(s) that exhibits changes in the derivative absorbance as a function of analyte concentration and which also exhibits a low variability in the derivative of absorbance between apparatus for a given analyte concentration for each relevant wavelength.

20

In another aspect of the invention, optimization comprises including absorbance and wavelength variability of more than one apparatus as follows:

25

(i) Obtaining a first set of interpolated absorbance measurements of the same primary calibration samples from the first apparatus and one or more second apparatus; and

(ii) Including all said interpolated absorbance measurements (step (i)) in the primary calibration set to select the optimal wavelength(s) to be

30

used to develop the primary calibration algorithm, by the process of step-wise multiple linear regression.

Only the data from the First Apparatus should be used to develop the
5 primary calibration algorithm at said optimized wavelength(s).

Photometric correction or absorbance adjustment of said interpolated absorbance may or may not be required.

10 H) SYNTHETIC CALIBRATION FLUIDS

Different primary calibration algorithms should be developed for an analyte in any one type of vessel, including, for example, labeled tube, unlabeled tube, pipette tip, tubing, translucent container or transparent container. Also, a calibration algorithm for an analyte may be developed for a
15 combination of several different types of vessels. As is clear the calibrators will be measured in a first apparatus with the calibrator in a vessel. According to a preferred embodiment of the method of the invention, measurements of the calibrators in a second apparatus are conducted with the calibrators in the same type of container as used for the first apparatus.

20

A further aspect of consideration with respect to calibrators is that the calibrator used for measurements in the first apparatus and the second apparatus be from the same batch. In a preferred embodiment, a large batch of calibrators is prepared so that there will be a significant period of time
25 before the batch expires or is depleted. In any event, at some time a new batch of calibrators will be required and that time it is preferable that a further primary calibration be conducted. This primary calibration may be conducted on the same first apparatus as used in respect of the first batch of calibrators or a different first apparatus may be used. In addition, different calibrators
30 may be chosen and the batch may therefore include different calibrators.

Once the new batch of calibrators is chosen and prepared, the primary calibration as described above is conducted in respect of the first apparatus and as described above, absorbance measurements of the synthetic calibrators are taken, and this information is provided in the calibration
5 package which will be used for calibration and recalibration of apparatus.

The methods just described generally require the use of a first apparatus or a new primary calibration. However, it is also possible to assign absorbance values to a new batch of calibrators without the use of the first
10 apparatus that is in control, and without conducting a cumbersome primary calibration. Alternative solutions to preparing a second batch or lot of calibrators are described under the title "Assigning Absorbances to a second Batch of Calibrators."

15 The use of such calibrators is preferred to primary calibration because of the following: a much smaller number of synthetic calibrators is required; synthetic calibrators are ready to use - materials for a primary calibration may not always be so readily available and cannot be stored for more than one day; synthetic calibrators are stable at room temperature and have a long
20 shelf life. The materials used in a primary calibration do not have either of these attributes. Primary calibration algorithms require a skilled user, and are usually developed off-line. Finally, with a calibration package, synthetic calibrators are relatively easy to use.

25 A preferred formulation of calibrators used in Example 1 is a combination of amaranth, phenol red, copper sulfate, and toluidine blue O, in 100 millimoles per liter acetate buffer, pH 3 to 4.

30 I) ASSIGNING ABSORBANCES TO A SECOND BATCH OF CALIBRATORS

According to another aspect of the invention where the calibrators are synthetic fluids, there are provided methods for assigning absorbances to the calibrators from new lots or batches, before or after the old lot or batch expires. Accordingly, the present invention provides methods for assigning
5 absorbance values to a second batch of synthetic calibrators on an apparatus comprising:

- 1) A method for assigning absorbance values to sets of synthetic calibration fluids from a second batch on a first apparatus, comprising:
10 (i) Testing whether said first apparatus is in control;
 (ii) If the test in step (i) is positive, then obtaining the absorbances measured on the first apparatus for at least one wavelength of the standard set of wavelengths, using at least one set of calibrators from the second batch; and
15 (iii) Assigning the mapped absorbances obtained in step (ii) to sets of calibrators from the second batch.

- 2) A method for assigning absorbance values to sets of a second batch of calibrators on a Second Apparatus having a previously established first set
20 of first linear regression equations for absorbance measurements using sets of calibrators from a first batch, comprising:
 (i) Testing whether said Second Apparatus is in control;
 (ii) If the test in step (i) is positive, then
 (a) retaining the previously established first set of first linear
25 regression equations;
 (iii) Obtaining absorbance measurements on the second apparatus for at least one wavelength of the standard set of wavelengths, using at least one set of calibrators from the second batch;
 (iv) Adjusting the mapped absorbances obtained in step (iii) for said
30 at least one set of calibrators from the second batch using the retained,

previously established first set of First linear regression equations from step (ii)(a); and

(v) Assigning the adjusted absorbances in step (iv) to sets of calibrators from the second batch.

5

3) A method for assigning absorbance values to sets of a second batch of calibrators on an apparatus having a previously established first set of first linear regression equations for absorbance measurements using sets of calibrators from a first batch, comprising:

10

(i) Testing whether said apparatus is in control;

(ii) If the test in step (i) is negative, then

(a) obtaining absorbance measurements on the apparatus for at least one wavelength of the standard set of wavelengths, using at least one set of calibrators from the first batch, and

15

(b) establishing a second set of first linear regression equations for the mapped absorbances obtained in step (ii)(a);

(iii) Obtaining absorbance measurements on the Apparatus for at least one wavelength of the Standard Set of Wavelengths, using at least one set of calibrators from the second batch;

20

(iv) Adjusting the mapped absorbances obtained in step (iii) for said at least one set of calibrators from the second batch using the retained, previously established set of first linear regression equations from step (ii)(a); and

(v) Assigning the adjusted absorbances in step (iv) to sets of
25 calibrators from the second batch.

J) CALIBRATION PACKAGE

The essential components of a Calibration Package comprises of:

(i) At least one primary calibration algorithm;

30

(ii) At least two calibrators; and

(iii) Assigned absorbances for said calibrators.

The assigned absorbances for calibrators are preferably obtained from the first apparatus when the apparatus is in control. By definition of first apparatus, the primary calibration algorithm(s) is or are developed for the first apparatus. Therefore, the assigned absorbances of calibrators and the primary calibration algorithm(s) should preferably originate from the same apparatus. During recalibration, if the calibrators' assigned absorbances do not originate from the same apparatus from which the primary calibration algorithm(s) originated, there exists a risk of having a mismatch between the calibrators' assigned absorbances and the primary calibration algorithm(s). In order to mitigate this risk, it is preferred that the appropriate primary calibration algorithm(s) be reincorporated in the apparatus. It becomes obvious that this preferred form of recalibration is no different from the process of calibration. Therefore, in the preferred embodiment, calibration is a safer approach than recalibration. In other words, calibration is preferred to recalibration when a calibration package is available.

Other software may be designed to prepare the software component of calibration packages, although any other means by which the primary calibration algorithms and assigned absorbances of calibrators (tested on the same first apparatus used to generate the primary calibration algorithm when said first apparatus is in control) may be transferred are within the scope of the present invention. In one embodiment of the invention, the at least one primary calibration algorithm and the assigned absorbances of calibrators are stored on a diskette although any other means of information transfer is contemplated, including CD-ROM, e-mail, internet information packages. These approaches will carry the assigned absorbances of the calibrations and also carry the primary calibration algorithm(s) from the first apparatus to the second apparatus. Preferably, calibration packages should contain

calibrators, to enable users to calibrate other apparatus when necessary, with minimum effort.

A calibration package will most preferably contain the following pieces
5 of information: the identity or the serial number of the first apparatus (i.e., the
apparatus used to develop the primary calibration algorithms); the calibration
algorithms and the corresponding analyte for each, developed on the first
apparatus; the vessel used for primary calibration and used for assigning
absorbances to the calibrators using the first apparatus must be checked
10 from a list of sample containers e.g., tips, tubes, tubing; the lot number of the
calibration set; the assigned absorbance values for each calibrator, at all the
relevant wavelengths; the analytes for which each calibrator from a set may
be applied. In a preferred embodiment, each calibrator is tested 10 times and
the average of 10 measurements for each wavelength may be used, instead
15 of a single measurement.

Although the description uses interferents as an example, it is obvious
that the invention can be applied to other analytes in other samples. Also,
although synthetic calibrators are used as examples, it should be understood
20 that other calibrators may be used with this invention.

The following non-limiting examples are illustrative of the present
invention:

25

EXAMPLES

Example 1

The absorbance spectra of four calibrators when tested on a first
apparatus are shown in Figure 1; the absorbance spectra obtained from a
second apparatus are shown in Figure 2. The first linear regression plots and
30 equations for standard wavelengths 500nm, 550nm, 600nm, 650nm, 700nm

and 750nm are shown in Figures 3, 4, 5, 6, 7, and 8 respectively. Table 1 shows the y-intercepts and slopes for the wavelengths chosen as examples.

Table 1

5

Wavelength (nm)	y-intercept	Slope	R ²
500	-0.796	1.464	0.99995
550	-0.801	1.465	0.99995
600	-0.664	1.464	0.99995
650	-0.744	1.466	0.99996
700	-0.732	1.466	0.99997
750	-0.785	1.467	0.99997

The slopes are very similar for the wavelength example shown in Table 1, but there are more significant differences for the y-intercepts. These numbers are very different from apparatus to apparatus. The large R² values (very
10 close to 1) indicates the high correlation between the absorbances of the two apparatuses, and hence the reliability of the photometric corrections.

The y-intercepts and slopes in Table 1 were derived from the first linear regression Plots of absorbances for the calibrators obtained on the first
15 apparatus (x-axis) and second apparatus (y-axis).

The equation for wavelength 600 nm is $y = 1.464x - 0.664$, or in a more general form.

20

$$\text{Absorbance}_{\text{Second}} = 1.464 \times \text{Absorbance}_{\text{First}} - 0.664$$

Let us say that the absorbance for a particular sample at 600 nm on a Second Apparatus was 1.500. Before the calibration algorithm from the first apparatus (primary calibration algorithm) for any analyte can be applied to the absorbances measured on second apparatus, the following adjustments must
5 be made:

$$\begin{aligned}\text{Adjusted absorbance} &= (\text{Measured absorbance} - \{-0.664\})/1.464 \\ &= (\text{Measured absorbance} + 0.664)/1.464 \\ &= (1.500 + 0.664)/1.464 \\ 10 &= 1.478\end{aligned}$$

Where -0.664 is a y-intercept and 1.464 is the slope from table above, for 600 nm. All wavelengths involved in the calibration algorithms must be adjusted similarly.

15

Quality control must be tested before and after calibrators are tested on the first apparatus when preparing a calibration package, for assurance that the first apparatus is performing precisely as it was when the calibrator algorithms were developed. In order for a second apparatus to use the
20 primary calibration algorithms derived for a first apparatus, preferably the following two items are required:

- 1). A table like Table 1 which includes at least the wavelengths used in the primary calibration algorithm(s), although all wavelengths may be
25 incorporated;
- 2). A calibration set.

Example 2

For illustration, a primary Hb calibration algorithm was developed for
30 apparatus #1, (a "First Apparatus"), and the same calibration algorithm was

applied to data generated from five other apparatuses (#2, #3, #4, #5 and #6, all of which are second apparatus). In this example, calibration between instruments is illustrated using the first derivative of the absorbance spectra.

5 The absorbance spectra for the same sample was shown for only three of the apparatus for clarity (#2, #4 and #6); they demonstrated the full range of variability in the absorbance spectra for a standard set of wavelengths for the six apparatus used, without photometric correction (Figure 9). When the first derivative of absorbance was calculated and plotted (Figure 10), most of
10 the variability disappeared, especially at wavelengths 593nm and 608nm, which were used in the calibration algorithm for hemoglobin, shown below:

$$\text{g/L Hb} = -0.22 - 75.96 \times (\text{1st derivative of absorbance at 593nm}) + 81.45 \times (\text{1st derivative of absorbance at 608nm})$$

15

Figure 10 is a graphic representation of the first derivative of absorbance spectra of the same synthetic fluid tested on 3 different apparatus (#2, #4, #6) using a standard set of wavelengths. The wavelengths identified as 593nm and 608nm were used in the calibration algorithm for hemoglobin. A
20 third wavelength, 537nm, is shown to illustrate an area with significant differences in both the absorbances and first derivatives of absorbance.

Scatter plots of the predicted hemoglobin on instrument #1 (x-axis) vs the five other instruments (y-axis) are shown in Figures 11 to 15.

25

Figure 11 is a scatter plot of the predicted Hb in g/L using the same calibration equation {g/L Hb = - 0.22 - 75.96 X (1st derivative of absorbance at 593nm) + 81.45 X (1st derivative of absorbance at 608nm)} applied to data collected from instrument #1 (x-axis) and instrument #2 (y-axis)

30

Figure 12 is a scatter plot of the predicted Hb in g/L using the same calibration equation $\{g/L\ Hb = -0.22 - 75.96 X (1st\ derivative\ of\ absorbance\ at\ 593nm) + 81.45 X (1st\ derivative\ of\ absorbance\ at\ 608nm)\}$ applied to data collected from instrument #1 (x-axis) and instrument #3 (y-axis)

5

Figure 13 is a scatter plot of the predicted Hb in g/L using the same calibration equation $\{g/L\ Hb = -0.22 - 75.96 X (1st\ derivative\ of\ absorbance\ at\ 593nm) + 81.45 X (1st\ derivative\ of\ absorbance\ at\ 608nm)\}$ applied to data collected from instrument #1 (x-axis) and instrument #4 (y-axis)

10

Figure 14 is a scatter plot of the predicted Hb in g/L using the same calibration equation $\{g/L\ Hb = -0.22 - 75.96 X (1st\ derivative\ of\ absorbance\ at\ 593nm) + 81.45 X (1st\ derivative\ of\ absorbance\ at\ 608nm)\}$ applied to data collected from instrument #1 (x-axis) and instrument #5 (y-axis)

15

Figure 15 is a scatter plot of the predicted Hb in g/L using the same calibration equation $\{g/L\ Hb = -0.22 - 75.96 X (1st\ derivative\ of\ absorbance\ at\ 593nm) + 81.45 X (1st\ derivative\ of\ absorbance\ at\ 608nm)\}$ applied to data collected from instrument #1 (x-axis) and instrument #6 (y-axis)

20

These data clearly demonstrate that for the calibration algorithm of Hb set out in this example, no photometric correction was necessary because the two wavelengths chosen, demonstrated very little variability in the first derivative of absorbance that was used in the calibration algorithm.

25

Example 3

Upon examination of the spectra in Figure 9, it may be seen that photometric correction would be necessary if the raw absorbances at any wavelength were used in the calibration algorithm. Upon examination of the first derivative of absorbance spectra of the same synthetic fluid tested on the

30

same 3 different apparatus (#2, #4, #6) in Figure 9, it is observed that there are no differences in the apparatus at 593nm and 608nm, using a standard set of wavelengths. It can be appreciated by those skilled in the art that the requirement of photometric correction depends on the required accuracy of the predicted analyte.

Example 4

Upon examination of the spectra in Figure 10, it is observed that photometric correction would be necessary if the first derivative of absorbance at a wavelength like 537nm was a major component in the primary calibration algorithm. It is also obvious that the requirement of photometric correction depends on the required prediction accuracy of the analyte, and the number of wavelengths like 593nm used in the calibration algorithm. If PLS analysis or PCA, which use the full spectrum or continuous segments of the spectrum (raw absorbance or an order of derivative of absorbance) is used, there may be a greater need to perform photometric correction.

Example 5

Figure 16 is a graphic representation of the second derivative of absorbance spectra of the same synthetic fluid tested on 3 different apparatuses (#2, #4, #6) using a standard set of wavelengths. The wavelengths identified as 593nm and 608nm were used in the calibration algorithm for hemoglobin when the first derivative of absorbance was used. For the purpose of calibration algorithm transfer one can visually observe the variability in the second derivative of absorbance; if the second derivative of absorbance was used to generate a calibration algorithm for hemoglobin, it seems obvious that 630nm (shown) would be preferred to 593nm. It should be noted that the best wavelength(s) used for calibration algorithm transfer (i.e., minimize inter-apparatus variability in predicted analyte concentration)

are not necessarily the best wavelength(s) that produce(s) accurate hemoglobin predictions.

It is very clear from Figure 9 (Absorbance) and Figure 10 that the first derivative of absorbance removes most of the spectral variability between the three apparatuses. However, as higher derivatives are used, the number of peaks in the spectrum increases, and therefore the peaks become sharper. To illustrate, compare the number of peaks and their sharpness in Figure 16 (second derivative absorbance) with Figure 15 (first derivative absorbance). This is seen even more so in the third derivative (Figure 19). Beyond the second derivative, the number of peaks do not change, but they become progressively sharper.

Example 6

A manufacturer calibrates spectrophotometers by using a single laser projected on the same pixel of each linear diode array detector, and uses a predetermined pixel dispersion. Therefore the apparatus using these spectrophotometers all have the same wavelength calibration table. Assume that three such apparatus are #2, #4, and #6 used to generate the absorbance spectra shown in Figure 9. Furthermore, assume that there is an error in the wavelength calibration of apparatus # 6. In order to simulate this error, the absorbances of the sample tested on apparatus #6, are shifted by 2 pixels. Figure 17 is a graphic representation of the second derivative of absorbance spectra of the same sample tested on 3 different apparatuses using a standard set of wavelength, and shifting the second derivative of absorbances for apparatus #6 to the right by 2 pixels. Figure 18 is a graphic representation of the first derivative of absorbance spectra of the same sample tested on 3 different apparatuses (#2, #4, #6) using a standard set of wavelength, and shifting the first derivative of absorbances for apparatus #6 to the right by 2 pixels. Although the derivative of absorbance may remove

some of the variability in absorbance as explained in Example 2, the impact of wavelength inaccuracy on the derivative of absorbance is demonstrated in Figure 17 and Figure 18, for the second derivative of absorbance and first derivative of absorbance respectively. This example demonstrates the importance of wavelength accuracy between apparatus. The magnitude of the wavelength error that is acceptable depends on the required accuracy of the analyte measurement. A preferred method to deal with wavelength inaccuracy is to use two wavelengths for wavelength calibration, whereby the pixel dispersion can be calculated for each apparatus; generate a wavelength calibration table for each apparatus, and map the measured absorbances to a standard set of wavelengths. It is also preferred that subsequent to the mapping of absorbances, that a Calibration Set be used to perform photometric correction for the relevant wavelengths.

While the present invention has been described with reference to what are presently considered to be preferred examples, it is to be understood that the invention is not limited to the disclosed examples. To the contrary, the invention is intended to cover various modifications and equivalents included within the spirit and scope of the appended claims.

20

The above description is not intended to limit the claimed invention in any manner, furthermore, the discussed combination of features might not be absolutely necessary for the inventive solution.

25 All publications, patents and patent applications referred to herein are incorporated by reference in their entirety to the same extent as if each individual publication, patent or patent application was specifically and individually indicated to be incorporated by reference in its entirety.